

Interferencije svojstvene kvantitativnim imunokemijskim metodama

Interferences in quantitative immunochemical methods

Slavica Dodig

Odjel za kliničko laboratorijsku dijagnostiku, Dječja bolnica Srebrnjak, Referentni centar ministarstva zdravstva za kliničku alergologiju djece, Zagreb
Department of Clinical Laboratory Diagnosis, Srebrnjak Children's Hospital, Reference Center for Clinical Allergology in Children of the Ministry of Health and Social Welfare, Zagreb, Croatia

Sažetak

Antitijelo koje se u imunokemijskim metodama koristi kao reagens otkriva ciljni analit (antigen). Iako je nekovalentna veza između analita i komplementarnog antitijela specifična, moguće su lažno pozitivne ili negativne interferencije. Neke su interferencije slične interferencijama kod kemijskih analiza, a neke su svojstvene samo imunokemijskim analizama. Na interferencije u imunokemijskim metodama treba pomisliti ako se dobije neprihvativ rezultat, ako postoji nelinearnost prilikom razrijedjivanja, ako nema podudarnosti s ostalim nalazima, odnosno kliničkim podacima, ako se različitim imunokemijskim metodama dobiju značajno različiti rezultati određivanja istog analita. U ovom će radu biti opisane neke od mogućih interferencija: 1) križna reaktivnost s endogenim i egzogenim substancijama koje nemaju strukturu antitijela; 2) križna reaktivnost s endogenim i egzogenim supstancama koje imaju strukturu antitijela; 3) prozonski učinak – hook efekt; i 4) utjecaj matriksa. Poznavanjem i prepoznavanjem interferencija u imunokemijskim analizama mogu se izbjegići moguće neželjene posljedice: pogreške u dijagnozi, liječenju i praćenju uspješnosti liječenja, nepotrebna dodatna laboratorijska istraživanja, nepotrebna terapija.

Ključne riječi: imunokemijske metode; interferencije; križna reaktivnost; prozonski učinak

Abstract

In the immunoassays, an antibody used as a reagent, detects an analyte (antigen) of interest. Although the noncovalent bound between analyte and complementary antibody is specific, false-positive and false-negative interferences are possible. Some interferences are similar to those in chemical analyses and some are typical only for immunoassays. One should suspect interferences in following cases: upon receiving an unacceptable result, if there is non-linearity during dilution, if there is no agreement with other test results or clinical data, if different immunoassays in determination of the same analyte provide significantly different results. This paper reviews some of the possible interferences: 1) cross-reactivity with endogenous and exogenous non antibody-structured substances; 2) cross-reactivity with endogenous and exogenous antibody-structure substances; 3) the hook effect; and 4) the matrix effect. By knowing and recognizing interferences in immunoassays, one can avoid possible undesired consequences: diagnostic errors, treatment and monitoring of its efficacy, unnecessary additional laboratory testing, unnecessary therapy.

Key words: immunoassays; interferences; cross-reactivity; prozone effect

Pristiglo: 22. kolovoza 2008.

Received: August 22, 2008

Prihvaćeno: 8. prosinca 2008.

Accepted: December 8, 2008

Uvod

Razvoj imunokemijskih metoda, osobito zadnjih tridesetak godina, revolucionirao je laboratorijsku medicinu. Primjena novih biljega završne reakcije, novih oblika testova, automatizacija, ponovljivost, brzina izvođenja i dos-

Introduction

The development of immunoassays has revolutionized laboratory medicine, especially during the last 30 years. The implementation of new endpoint tracers, new assay formats, automation, reproducibility, duration time of

tupnost analiza pridonijeli su da imunokemijske metode postanu svakodnevna praksa. Osnovno svojstvo svih imunokemijskih metoda – od imunoprecipitacijskih do biočip metoda – jest da reagens, kojim se otkriva ili kvantitativno određuje ciljni analit (antigen), sadrži antitijelo. Iako je nekovalentna veza između analita i komplementarnog antitijela specifična, moguće su brojne interferencije (Slika 1.) koje uzrokuju dobivanje lažno povećanih (pozitivna interferencija) (1-3) ili lažno smanjenih rezultata (negativna interferencija) (4,5). U svakodnevnom radu nužno je misliti na uvijek prisutne predvidljive i uvijek moguće nepredvidljive i neprepoznatljive interferencije (6). Jedan od najdrastičnijih primjera pogreške u medicinskoj praksi jest primjer lažno-pozitivnog nalaza humanog korionskog gonadotropina (hCG), opisan kod 22-godišnje žene koja je, zbog neprepoznate interferencije heterofilnih antitijela te stoga trajno lažno-pozitivnog nalaza hCG, podvrgnuta nepotrebnim medicinskim zahvatima: kemoterapiji, histerektomiji i segmentalnoj plućnoj resekciji (7). Taj je slučaj dobio pozornost javnih glasila (odšteta 16 milijuna USD). Međutim, stručno-znanstvena publicistika opisuje slične slučajeve (8,9).

Interferencije bi se mogle definirati kao učinak tvari prisutnih u analitičkom sustavu koje uzrokuje promjenu izmjerene vrijednosti u odnosu na pravu vrijednost (10). Neke su interferencije svojstvene imunokemijskim metodama. Posebne preanalitičke i analitičke interferencije utječu na kliničko vrednovanje imunokemijskih nalaza u usporedbi s ostalim kemijskim analizama. Zbog osobitih svojstava imunokemijskih metoda (križna reaktivnost antitijela, specifičnost, ograničenja analitičke osjetljivosti, utjecaj matriksa, itd.) može doći do nesklada laboratorijskih nalaza.

Proizvođači reagensa za imunokemijske analize dužni su upozoriti na te interferencije, a neke su obično naznačene u uputama za izvođenje analitičkog postupka. Interferencije koje ovise o analitu odnose se na interakciju tvari prisutnih u biološkom uzorku i jedne (1,11-16) ili više komponenata iz reagensa (17,18). Učinak interferencije obično ovisi o koncentraciji interferenta (12,19). Interferencije koje ovise o analitu podrazumijevaju spojeve koji su strukturno slični analitu. Zbog toga ti spojevi reagiraju križno s antitijelom, ili ostalim proteinima u uzorku, primjerice autoantitijelima prema analitu, heterofilnim antitijelima, humanim anti-animalnim antitijelima (12). Najčešće opisivane interferencije su one pri određivanju hormona (5,20-24), tumorskih biljega (25), lijekova i metabolita (26,27), troponina (29-31), te pri serološkim analizama (32-33). Opisano je da antitijela prema analitu (autoantitijela) uzrokuju interferencije za brojne analite, primjerice za tiroroidne hormone (ukupne i slobodne oblike), tireoglobulin, prolaktin (makropolaktinemija može imati kao posljedicu hiperprolaktinemiju, a da ne postoji bolesti hipofize), testosteron, izulin (12). Antitiroидna autoantitijela su

assay and availability of analyses have contributed for immunoassays to become everyday practice. The main characteristic of all immunoassays – from immunoprecipitation to biochip assays – is that the reagent that discovers or quantifies the target analyte (antigen) contains the antibody. Despite the specificity of the noncovalent bond between analyte and complementary antibody, numerous interferences (Figure 1) are possible, and can cause false increase (positive interference) (1-3) or false decrease of measured result (negative interference) (4,5). In every research, it is necessary to think about always present predictable and always possible unpredictable and unrecognizable interferences (6). One of the most drastic examples of error in medical practice is the case of false positive chorionic gonadotropin (hCG) test result, described with 22 year old women who underwent, due to unrecognized interference of heterophilic antibodies followed by permanent false positive hCG test result, unnecessary medical interventions: chemotherapy, hysterectomy and segmental lungs resection (7). That case won a lot of public attention (16 million USD damage was paid). Moreover, scientific literature offers similar cases (8,9). Interferences may be defined as the effect of substances present in an analytical system which causes deviation of the measured value from the true value (10). Some interferences are typical for immunochemical methods (11). The special pre-analytical and analytical interferences have their impacts on the clinical evaluation of immunochemical findings, compared to other chemistry methodologies. Due to the specific features of the immunoassay technique (cross-reactivity of the antibodies, specificity, technology-dependent sensitivity limits, the matrix effect, etc.), which might cause misleading laboratory report. Manufacturers of reagents for immunoassays are obliged to warn against these interferences, and some of them can usually be found in instructions for analytic procedure. Analyte-dependent interferences relate to interaction of substances present in biological sample and one (1,11-16) or more reagent components (17,18). The interference effect usually depends on concentration of interfering substance (12,19). These analyte-dependent interferences include compounds that are structurally similar with analyte. Therefore, they cross-react with the antibody or other proteins in the sample, e.g. autoanalyte antibodies, heterophile antibodies, human anti-animal antibodies or rheumatoid factors (12). The most frequently described interferences are the ones occurring during hormone determination (5,20-24), tumor markers (25), drugs and metabolites (26-28.), troponin (29-31) and during serological analyses (32,33). Autoanalyte antibodies (autoantibodies) have been described to cause interferences for a number of analytes, e.g. thyroid hormones (total and free forms), thyroglobulin, prolactin (macroprolactinemia can result in hyperprolactinaemia without pituitary

nađena u bolesnika s Gravesovom bolesti, Hashimotovim tireoiditisom, hipertireoidizmom nakon liječenja, u bolesnika s gušom, karcinomom ili ne-tireoidnim autoimunim stanjima. Prisustvo reumatodnog faktora u serumu može uzrokovati lažno povećane koncentracije troponina, kao i interferirati kod ispitivanja funkcije štitnjače. Zbog steričkog blokiranja reakcije analit-antitijelo, paraprotein može također interferirati u imunokemijskim analizama. Na interferencije u imunokemijskim metodama treba pomisliti ako se dobije neprihvatljiv rezultat, ako postoji ne-linearnost prilikom razrjeđivanja, ako nema podudarnosti s ostalim nalazima, odnosno kliničkim podacima, ako se različitim imunokemijskim metodama dobiju značajno različiti rezultati određivanja istog analita. Nepoznavanje i neprepoznavanje interferencija može imati kao posljediku pogreške u dijagnozi, liječenju i praćenju uspješnosti liječenja, nepotrebna dodatna laboratorijska istraživanja, nepotrebnu terapiju (kod lažno smanjene koncentracije analita bolesnik se može predozirati). Većina je interferencija svojstvena svim oblicima imunokemijsih metoda (Tablica 1.), a neke se interferencije odnose na pojedine metode.

Interferirajuća antitijela mogu utjecati na sve vrste imunokemijskih analiza, ali su najčešća u saturacijskim analizama. Razlog tome jest taj što saturacijske analize podrazumijevaju suvišak obaju antitijela (primarnog i obilježenog). Njihova je koncentracija veća od uobičajene koncentracije analita, pa se reakcija odvija vrlo brzo u uvjetima velike analitičke osjetljivosti. Bilo koje antitijelo iz seruma koje ima i neznatan afinitet prema primarnom i obilježenom antitijelu može s njima stvarati mjerljivi kompleks. Sva antitijela seruma koja su dovoljno velika da mogu vezati istodobno dva antitijela iz reagensa, na kraju daju mjerljivi signal. Jedan od čestih primjera interferenta su idiotipska antitijela, primjerice reumatodini faktori koji sadrže križno reaktivne idiotope (25).

U ovom će radu biti pobliže opisane neke od mogućih interferencija svojstvenih imunokemijskim metodama, poglavito one koje bi klinički biokemičar morao poznavati: 1) križna reaktivnost s endogenim i egzogenim supstancama koje nemaju strukturu antitijela; 2) križna reak-

disease), testosterone, insulin (12). Antithyroid autoantibodies have been found in patients with Graves' disease, Hashimoto's thyroiditis, hyperthyroidism after treatment, goiter, carcinoma or non-thyroid autoimmune conditions. The presence of rheumatoid factors in serum can cause falsely elevated values in troponin assays as well as in thyroid function methods. Also, paraprotein can interfere in immunoassay by sterical blocking analyte-antibody reaction.

One should suspect interferences in immunoassays upon receiving an unacceptable result, if there is non-linearity during dilution, if there is no agreement with other test results or clinical data, if different immunoassays in determination of the same analyte provide significantly different results. Unawareness and non-recognition of interferences could lead to diagnostic errors, inadequate treatment and monitoring of its efficacy, unnecessary laboratory tests, unnecessary therapy (falsely low analyte concentration can lead to patient overdose). Most of the interferences are typical for all immunoassays (Table 1) and some relate to single methods.

Interfering antibodies can affect all types of immunochemical analyses, but they are most frequently present in saturating analyses. That is because in saturating analyses we have an excess of both antibodies (the capture and the tracer one). Their concentration is higher than the usual analyte concentration and the reaction occurs very fast in conditions of high analytic sensitivity. Any serum antibody having only slight affinity to the capture and the tracer antibody can together with them create a measurable complex. All serum antibodies big enough to bind simultaneously two antibodies from the reagent in the endpoint, provide a measurable signal. One of the common examples of interfering substances are idiopathic antibodies, e.g. rheumatoid factors containing cross-reactive idiotopes (25).

This paper gives a close review of some of the possible interferences typical for immunoassays, especially of those of greater importance for a clinical biochemist: 1) cross-reactivity with endogenous and exogenous non antibody-structured substances; 2) cross-reactivity with endo-

TABLICA 1. Interferencije kod pojedinih imunokemijskih metoda**TABLE 1.** Interferences in particular immunoassays

Interference	Methods
Cross-reactivity	All, but mostly competitive
Prozone effect (hook effect)	Nephelometric, turbidimetric, saturating
Matrix	All
Antibodies	All, but mostly saturating

tivnost s endogenim i egzogenim supstancama koje imaju strukturu antitijela; 3) prozonski učinak - *hook* efekt; i 4) utjecaj matriksa.

Križna reaktivnost s endogenim i egzogenim supstancama koje nemaju strukturu antitijela

Križna je reaktivnost najčešća interferencija u imunokemijskoj, ali najčešće u kompetitivnim metodama. Radi se o nespecifičnom utjecaju tvari u uzorku, koja je strukturno slična analitu (ima jednake ili slične epitope kao analit) te nadmeće se za vezno mjesto na antitijelu (34). Stupanj interferencije uzrokovane križnom reaktivnošću ovisi o tri čimbenika: specifičnosti antitijela, obliku testa i pripremi uzorka (35). Najčešći su primjeri prilikom određivanja koncentracije hormona, lijekova, specifičnog IgE prema alergenima. Hormoni TSH (tireotropin, engl. *thyroid-stimulating hormone*), LH (luteinizirajući hormon; engl. *luteinizing hormone*), FSH (folikulo-stimulirajući hormon; engl. *follicle-stimulating hormone*) i hCG (humani korionski gonadotropin; engl. *human chorionic gonadotropin*) imaju analogan α -lanac, a β -lanac određuje specifičnost pojedinog hormona – stoga treba odabrati metodu koja će specifičnim antitijelima moći prepoznati različite epitope (36). Drugi je primjer steroidnih hormona, koji imaju jednaku ciklopentanoperhidrofenantrensku strukturu (37,38). Nadalje, prostatični specifični antigen (engl. *prostate specific antigen*, PSA) – postoji u nekoliko oblika [ukupni PSA, slobodni fPSA, prekursor proPSA, a u novije vrijeme i antigen ranog rasta karcinoma prostate (engl. *early prostate cancer antigen*, EPCA)], koji su uzrok križnoj reaktivnosti i nedovoljnoj točnosti (39). Križna reaktivnost uzrokuje lažno povećane vrijednosti ispitivanog analita, ali, ovisno o obliku testa, može uzrokovati i lažno smanjene vrijednosti analita (Slika 1Ab).


Križnu reaktivnost može prouzročiti metabolit ili prekursor analita, primjerice konjugirani metaboliti kortizola pri određivanju kortizola u mokraći (38), ili istodobna primjena lijekova slične strukture (triciklički antidepresivi) (19). Poznat je problem križne reaktivnosti kod određivanja vitamina D ($1,25\text{-}[OH]_2D_3$) zbog moguće pozitivne interferencije $25\text{-}OH D_3$ (34). I u području alergologije poznate su interferencije (40) pri određivanju specifičnih IgE prema alergenima kravljeg mlijeka (41), alergenima grinja (42), plodovima mora (43), peluda i lateksa (44), epitela životinja (45), alergena otrova opnokrilaca (46). Opisana je također lažno povećana koncentracija IgE na alergene peluda biljaka, zbog prisustva IgE prema ugljikohidratnim determinantama monoglikozilirane alergenske molekule (47,48). U tom slučaju osoba, unatoč povećanoj koncentraciji IgE, nema simptoma alergijske bolesti. Razlog tome je taj, što IgE antitijela prema ugljikohidratnim determi-

genous and exogenous antibody-structured substances; 3) the hook effect; and 4) the matrix effect.

Cross-reactivity with endogenous and exogenous non antibody-structured substances

Cross-reactivity is the most common interference in immunoassays, but mostly in competitive ones. It is non-specific influence of substances in a sample that structurally resembles analyte (carries similar or the same epitopes like analyte) and competes for binding site on antibody (34). The interference grade caused by cross-reactivity depends on three factors: antibody specificity, method and sample preparation (35). The most common examples can be seen during determination of hormone concentration, drugs, allergene-specific IgE. Hormones TSH (thyroid-stimulating hormone), LH (luteinizing hormone) and hCG (human chorionic gonadotropin) carry analogue α -chain, and β -chain determines the specificity of the respective hormone – therefore, it is necessary to choose an assay which would be able to recognise different epitopes by using specific antibodies (36). The second example are steroid hormones, which have identical cyclopentanoperhydrophenanthrene structure (37,38). Furthermore, there is prostate specific antigen (PSA) which exists in several forms (total, PSA; free, fPSA; precursor, proPSA and newly also early prostate cancer antigen, EPCA) and causes cross-reactivity and insufficient accuracy (39). Cross-reactivity causes falsely elevated concentration of analyte, but depending on the test method falsely low values can also occur (Figure 1 Ab).

Cross-reactivity can be caused by metabolite or analyte precursor, e.g. conjugated cortisol metabolites by determining the urine cortisol (38) or simultaneous application of medications with similar molecular structure (tricyclic antidepressants) (19). The problem of cross-reactivity by vitamin D ($1,25\text{-}[OH]_2D_3$) determination due to possible positive interference of $25\text{-}OH D_3$ (34) is well known. In the area of allergology one can also find interferences (40) by determination of allergene-specific IgE to cow milk (41), mite allergens (42), seafood (43), pollen and latex (44), animal epithelium (45), hymenoptera sting venom allergens (46). False increased concentration of IgE to pollen allergens due to presence of IgE to carbohydrate determinant of the monoglycosylated allergenic molecule have also been reported (47,48). In this case a person despite increased IgE concentration shows no symptoms of an allergic disease. The reason for that is that IgE antibodies to carbohydrate determinants do not affect histamine release from basophilic granulocytes or mast cells. Cross-reactivity usually causes positive interference, but in some assays negative interference is also possible. So, for example in digoxin-determinating assay oleandrin



SLIKA 1. Različite interferencije u imunokemijskim analizama: Aa - analiza bez interferencije; Ab - križna reaktivnost interferenta s veznim antitijelom, rezultira lažno negativnim rezultatom; B - pozitivna interferencija: a - nespecifično vezanje detektorskog antitijela na krutu podlogu koja nije obložena blokirajućim agensom; b - premoščavanje s heterofilnim antitijelima odnosno HAMA; C - negativna interferencija; Ca - promjena steričke konformacije zbog vezanja interferirajućeg proteina uz Fc fragment detektorskog antitijela; Cb - prikrivanje epitopa proteinom iz uzorka

FIGURE 1. Different interferences in immunoassays: Aa - assay without any interference; Ab - cross-reactivity of an interfering substance with capture antibody, resulting with false negative result; B - positive interference: Ba - unspecific binding of labelled detector (tracer) antibody to a not blocked solid phase; Bb - "bridge" binding by heterophilic antibodies or HAMA, respectively; C - negative interference: Ca - change of sterical conformation after binding of interfering protein to Fc fragment of detector antibody Cb - masking of the epitope on analyte surface by a protein of the sample.

nantama ne utječe na oslobađanje histamina iz bazofilnih granulocita odnosno mastocita.

Križna reaktivnost obično uzrokuje pozitivnu interferenciju, ali je u nekim testovima moguća i negativna interferencija. Tako npr. oleandrin (srčani glikozid sličan digoksinu) u testu određivanja digoksina može interferirati na različite načine (49). Pri smanjenim koncentracijama digoksina, oleandrin može imati pozitivnu interferenciju, a pri povećanim koncentracijama digoksina negativnu. U eri transplantacije organa osobito je važno znati da imunokemijske metode za određivanje koncentracije imunosupresivnog lijeka ciklosporina A daju značajno veću koncentraciju nego referentna metoda HPLC (49). Križna reaktivnost opisana je i u metodama za probiranje na

(digoxin-like cardiac glycoside) can interfere in different ways (49). By decreased digoxin concentrations, oleandrin can have positive interference and by increased concentrations negative interference. In the time of organ transplantations it is important to know that immunoassays for determination of cyclosporine A concentration, an immunosuppressive drug give significantly higher concentration than referent HPLC method (49). Cross-reactivity has also been reported in methods for drug misuse screening (50, 51). Faster dissociation of an interfering substance than of an analyte during washing or separating free from captured analyte during analysis can be the cause for falsely low concentration of an analyte (52).

zlouporabu lijekova (50,51). Uzrok lažno negativnim interferencijama može biti brža disocijacija interferenta nego analita pri ispiranju ili odvajajanju slobodnog od vezanog analita tijekom analize (52).

U kompetitivnim metodama određivanja malih molekula (lijekovi), oba antitijela, primarno (vezno) i obilježeno (detektorsko), vežu se istodobno uz analit. Križnu je reaktivnost teško predvidjeti, stoga moramo biti svjesni njenog postojanja, pratiti stručno-znanstvenu literaturu i odabirati specifičnije metode.

Križna reaktivnost s endogenim i egzogenim supstancama koje imaju strukturu antitijela

Na imunokemijsku reakciju mogu utjecati antitijela prisutna u biološkom uzorku bolesnika ili antitijela reagensa (13,53). Biološki uzorak može sadržavati egzogena i endogena antitijela. Endogena antitijela prisutna su u oko 40% osoba (14), osobito onih koje su dobivale imunoterapiju s monoklonalnim antitijelima (54). U skupinu egzogenih antitijela ubrajaju se imunološki lijekovi. Iz te se skupine najviše opisuje interferencija nakon intravenske primjene Fab fragmenta antidigoksinskih antitijela (Fab fragment usmjeren prema antigenskoj determinantni digoksinu; dobiva se iz antidigoksinskih antitijela proizvedenih u ovci) (55,56). Mechanizam interferencije Fab fragmenta podrazumijeva različiti afinitet i specifičnost primarnih antitijela u pojedinim testovima. Opisana je interferencija ginsenga (imunoreaktivna tvar slična digoksinu) (57).

Dvije su vrste endogenih antitijela u serumu pacijenta. To su heterofilna antitijela (prirodna antitijela i autoantitijela) (58,59) i anti-animalna antitijela (engl. *human anti-animal antibodies*, HAAAs) (23). Iako se endogena antitijela razlikuju prema nekim svojstvima (60), interferiraju prema jednakom mehanizmu u saturacijskim analizama - stvaraju komplekse istodobno i s primarnim i obilježenim antitijelima reagensa te ih premoštavaju (Slika 1.). Heterofilna antitijela su multispecifična antitijela sintetizirana prema slabo definiranim antigenima. Humana anti-animalna antitijela su antitijela velike avidnosti, a sintetiziraju se prema dobro definiranim antigenima (16).

Interferencija heterofilnih antitijela većinom ima za posljedicu lažno povećane (Slika 1B.) rezultate (58,60-64), iako su opisani i lažno smanjeni rezultati (31,55,65-67) u slučajevima kad interferirajuće antitijelo stvara kompleks samo s jednim antitijelom iz reagensa. Pozitivna interferencija heterofilnih antitijela u sendvič metodama nastaje zbog toga što heterofilna antitijela premoštavaju primarno i obilježeno antitijelo (68). Negativna interferencija nastaje zbog vezanja heterofilnih antitijela izravno na primarno antitijelo, što onemogućuje vezanje analita. U nekim imunokemijskim metodama (ELISA, luminometrijske metode) reagensi sadrže životinjske proteine (govedi albumin i kazein) koji služe za blokiranje reaktivnih

In competitive immunoassays of small molecules (drugs), both, the capture (binding) and the tracer antibody (labelled detector), bind simultaneously to the analyte. Cross-reactivity is hard to predict, therefore, we must be aware of its existence, be up to date with scientific literature and choose more specific methods.

Cross-reactivity with endogenous and exogenous antibody-structured substances

Immunoreaction can be influenced by antibodies present in biological sample of a patient or antibodies from the reagent (13,53). Biological sample can contain exogenous and endogenous antibodies. Endogenous antibodies are presented in about 40% of patients (14), especially in those who were on immunotherapy with monoclonal antibodies (54). Immunological drugs belong to the group of exogenous antibodies. From this group the most commonly reported interference is the one upon intravenously applied Fab fragment from antidigoxin antibodies (Fab fragment is directed to antigen determinant of digoxin; Fab fragment comes from antidigoxin antibody produced in sheep) (55,56). Interference mechanism of the Fab fragment has different affinity and specificity of capture antibodies in some assays. It has also been reported about interference of ginseng (digoxin-like immuno-reactive component) (57).

There are two types of endogenous antibodies in patients' serum. Heterophilic antibodies (natural antibodies and autoantibodies) (58,59) and anti-animal antibodies (human anti-animal antibodies; HAAAs) (23). Although endogenous antibodies differ in some characteristic (60) they interfere according to identical mechanism in saturating (sandwich) analyses – they simultaneously create complexes with capture and tracer antibodies of the reagent – they "bridge" them (Figure 1). Heterophilic antibodies are multi-specific antibodies synthesised to very poorly defined antigens. Human anti-animal antibodies are antibodies of high avidity and are synthesised to well defined antigens (16).

Heterophilic antibody interference has falsely elevated (Figure 1B) results as a consequence (58,60-64), although falsely low values (31,55,65-67) have been reported in cases when the interfering antibody creates a complex with only one antibody from the reagent. Positive heterophilic antibody interference in sandwich capture assays occurs because heterophilic antibodies "bridge" the capture and the tracer antibody (68). Negative interference occurs due to binding of heterophilic antibodies directly to the capture antibody what disables binding of analytes. In some assays (ELISA, luminometric methods) reagents contain animal proteins (bovine albumin and casein) which block reactive sites on microtiter plates or polystyrenic micro substances. However, their interference ability can

mjesta na mikrotitarskim pločicama ili polistirenским mikročesticama. Međutim, dvojaka je mogućnost njihove interferencije: mogu uzrokovati lažno povećane rezultate, ali mogu izazvati i povećani pozadinski signal (engl. *background*), ako se heterofilna antitijela izravno vežu na njih (68). Heterofilna antitijela interferiraju i pri određivanju citokina metodom ELISA (69), čija je koncentracija u serumu vrlo mala (66,70). Kako bi se izbjegao utjecaj heterofilnih antitijela pri određivanju koncentracije citokina, uzorku se dodaje ne-imuni animalni serum (68). Moguće su interferencije i u bolesnika s monoklonalnim gamapatijama (69,71) i u bolesnika s prisutnim autoantitijelima (31,72). Najpoznatija anti-animalna antitijela su humana anti-mišja antitijela (engl. *human anti-mouse antibodies*, HAMA) (14,53,73,74). Mišja monoklonalna antitijela sve se više primjenjuju intravenski u dijagnostičke ili terapijske svrhe u onkologiji (75), alergologiji (76), autoimunim bolestima (77), a neki pacijenti prema njima sintetiziraju HAMA. HAMA mogu interferirati s mišjim monoklonalnim antitijelima, ako su ona sastavni dio reagensa. Oko 10% pacijenata ima heterofilna antitijela (12), a oko 40% osoba koje su intravenski primile mišja monoklonalna antitijela sintetizirat će HAMA. Pojavnost antianimalnih antitijela veća je u bolesnika s manjkom IgA (39% ih ima anti-kozja, a 18% ih ima HAMA), nego osoba s normalnom koncentracijom IgA (22). Neke osobe mogu sintetizirati anti-animalna antitijela nakon što su bile izložene ostalim životinjskim antigenima, primjerice u cjeplivima dobivenim u zečevima ili kokošima, antizmijskim otrovom dobivenim u konju, profesionalnim kontaktima s kućnim ljubimcima i ostalim životnjama epitela životinja (45).

Reagensi za imunometrijske analize obično sadrže serum ili neki drugi blokirajući agens, koji bi trebao smanjiti nespecifičnu interferenciju (20,53,78-81). Ako se očekuje interferencija, korisno je imati dodatni blokirajući agens kojim bi se tretirali uzorci (78). Ne postoji univerzalni blokirajući agens za sve analite i za sve metode, nego se mora primijeniti onaj agens koji je nakon validacije pokazao da najbolje reducira utjecaj heterofilnih antitijela za određeni analit (53).

Prozonski učinak (engl. hook effect)


Prozonski učinak, u literaturi poznat i kao *hook* (engl. *hook*, savijen, poput udice) efekt, temelji se na krivulji zasićenja antitijela antigenom (Slika 2.). Primarno, prozonski učinak ovisi o koncentraciji analita (34,82-85). Podrazumiјeva stanje njegova izrazitog suviška, koji zasiti sva vezna mjesta na antitijelu (86-89). Učinak nastaje uglavnom (ali ne isključivo), u metodama kod kojih se sve tri sastavnice (antigen, antitijelo, biljeg) inkubiraju istodobno (engl. *single step assay*) (90). Prozonski učinak ne postoji kod kompetitivnih imunokemijskih analiza. To znači da u reakciji ostaje višak analita koji nije ušao u sastav kompleksa

go in two directions: they can cause falsely increased results, but also induce increased background signal if the heterophilic antibodies directly bind to them (68). Heterophilic antibodies interfere also during cytokine determination by ELISA (69) whose serum concentration is very low (66,70). Heterophilic antibody interference in cytokine determination can be reduced by adding normal non-immune animal serum to the sample (68). Interferences in patients with monoclonal gammopathies (69,71) are also possible, as well as interferences in patients with present autoantibodies (31,72).

The most common anti-animal antibodies are human anti-mouse antibodies (HAMA) (14,53,73,74). Mouse monoclonal antibodies are increasingly applied intravenously in diagnostic or therapeutic purposes in oncology (75), allergology (76), autoimmune diseases (77), and some patients synthesise HAMA to them. HAMA can interfere with mouse monoclonal antibodies if they are a component of the reagent. Around 10% of patients carry heterophilic antibodies (12) and around 40% patients who intravenously received mouse monoclonal antibodies will synthesise HAMA. Prevalence of anti-animal antibodies is higher in patients with IgA deficiency (39% carry anti-goat antibodies and 18% carry HAMA) than in patients with normal IgA concentration (22). Some patients can synthesise anti-animal antibodies after exposure to other animal antigens, e.g. in vaccine produced in rabbits or chicken, in anti-snake venom produced in the horse, in professional exposure to pets and other animals (45). Reagents for saturating analyses usually contain serum or some other blocking agent that should diminish the non-specific interference (20,53,78-81). If interference is expected, it is useful to have an additional blocking agent to treat the samples (78). There is no universal blocking agent for all analytes and all methods. The agent that after validation showed as the best for reducing the effect of heterophilic antibodies for certain analyte must be applied (53).

The hook effect

The hook effect is based on the saturation curve of antibody with antigen (Figure 2). Primarily, the hook effect depends on analyte concentration (34,82-85). It implies the presence of huge excess of analyte which saturates all binding sites on antibody (86-89). The effect occurs mostly (but not exclusively) in assays where all three components (antigen, antibody and marker) incubate simultaneously (single step assay) (90). The hook effect does not occur in competitive immunoassays. That means that in reaction there is a surplus on analytes that did not penetrate to analyte-antibody complex compound. This results in falsely decreased value of the measured analyte which could even lie in the reference interval. The value of



SLIKA 2. Prozonski učinak - Izrazito povećana količina analita nadmašuje vezni kapacitet primarnog antitijela. To rezultira neodgovarajuće slabim signalom koji uzrokuje pogrešno smanjen ili normalan rezultat ("hooked" result) u pacijenta kod kojeg postoji izrazito povećana koncentracija analita u serumu.

FIGURE 2. The hook effect - An excessive amount of analyte overwhelms the binding capacity of the capture antibody. This results in an inappropriately low signal that causes erroneous low or normal result ("hooked" result) for a patient with an excessively elevated serum analyte concentration.

analit-antitijelo. Posljedica jest lažno smanjena vrijednost ispitivanog analita, koja može biti čak unutar referentnih intervala. Dobije se apsorpcija u post-zoni (silazna strana krivulje) čija je vrijednost jednaka vrijednosti apsorpcije u pre-/pro-zoni (uzlazna strana krivulje). U tom slučaju, reakcijska krivulja ima zvonolik oblik (engl. *bell-shaped curve*), odnosno savijena je poput udice (engl. *hook*) (34). Neki automatski analizatori imaju sustav za prepoznavanje suviška analita uz istodobno razrjeđivanje uzorka. Većina automatskih analizatora za područje kliničke kemije samo upozorava na nelinearnu reakciju, što je upozorenje da je uzorak potrebno razrijediti. Proizvođači reagensa za imunoturbidimetrijska određivanja smanjili su prozonski učinak uvođenjem lateks čestica kao nosača na kojima se odvija reakcija između analita (antigena) i antitijela. Kod kompetitivnih metoda prozonski je učinak otklonjen postupkom ispiranja (svremeni automatski analizatori imaju programirano ispiranje) nakon reakcije analita s primarnim antitijelom i dodavanja obilježenog antitijela (86-90). Proizvođači reagensa smanjuju prozonski učinak povećanjem količine primarnog i obilježenog antitijela odnosno smanjenjem količine uzorka potrebnog za analizu. Prozonski učinak česta je pojava u svakodnevnom radu u kliničkim laboratorijima i nikako se ne smije zanemariti. Postoji kod onih analita koji se u serumu mogu naći u izrazito širokom rasponu koncentracija, kao što su primjerice C-reaktivni protein (stostruko povećanje), antistreptoliziinska antitijela (deseterostruko povećanje), hormoni,

absorbance in the post-zone (down-side of the curve) is identical with the absorbance value in the pre-/pro-zone (up-side of the curve). In this case, the reaction curve is bell-shaped (bell-shaped curve) or hooked (34). Some automated analysers have a system for recognizing excess of an analyte while the sample is being simultaneously diluted. The most automated analysers used in clinical chemistry only warn about non-linear reaction what is sufficient to see if the sample needs to be diluted. Manufacturers of reagents for immuno-turbidimetric determinations have reduced the hook effect by introducing latex particles as carriers on which the reaction between analyte (antigen) and antibody takes place. In competitive assays the hook effect was eliminated by introducing a wash step (this wash step is programmed in current automated analysers) upon reaction of analyte with the capture antibody and addition of the tracer antibody (86-90). Reagents manufacturers reduce the hook effect by increasing the quantity of the capture and the tracer antibody and by reducing the quantity of samples required for the analysis.

The hook effect is common phenomenon in everyday work of a clinical laboratory and on no account should be neglected. It exists by analytes present in serum in extremely wide range of concentrations like C-reactive protein (100-fold increase), antistaphylococcal antibodies (10-fold increase), hormones (at 6-fold concentration increase) (hCG), IgE (>1000-fold), ferritin (100-fold increase), tu-

npr. hCG (kod šesterostrukog povećanja koncentracije), IgE (>1000 puta), feritin (stostruko povećanje), tumorski biljezi (osobito CA 19-9, PSA) (34). Na određivanje tumorskih biljeza uglavnom utječe moguće veliko povećanje koncentracije (>10.000 puta), koje postoji u bolesnika s izrazitim tumorskim rastom. Može se pojaviti kod prve obrade bolesnika. Granična vrijednost, kod koje se gubi linearnost turbidimetrijskih metoda, pomaknuta je prema većim vrijednostima uvođenjem lateks čestica kao nosača antitijela. Ako proizvođač reagensa nije naznačio graničnu koncentraciju analita iznad koje se pojavljuje prozonski učinak, medicinski biokemičar bi to morao ispitati, te unijeti podatak u priručnik o kvaliteti rada laboratorija. Mogućnost postojanja prozonskog učinka otkriva se određivanjem uzorka s izrazito velikom koncentracijom u nerazrijeđenom uzorku i u razrjeđenjima, 1:10 i 1:100 (34,91). Ako se u razrjeđenim uzorcima dobije veći rezultat nego u nerazrijeđenom uzorku, radi se o prozonskom učinku. Potom slijedi određivanje granične koncentracije koja se može pouzdano odrediti. Uzorak se mora razrjeđivati sve dok se rezultati dvaju različitim razrjeđenjima podudaraju (uzimajući u obzir faktor razrjeđenja). Ako se unaprijed očekuje izrazito povećana vrijednost ispitivnog analita koja će dati lažno smanjenu vrijednost, mogu se odmah pripremiti dva uzorka - nerazrijeđeni i razrjeđeni.

Učinak matriksa

Uzorak seruma, odnosno plazme, složena je smjesa lipida, proteina, ugljikohidrata, soli i vode. Zbroj interferencija svih sastavnica u uzorku (osim analita), koje utječu na mjerjenje ciljnog analita, poznat je pod nazivom „učinak matriksa“ (92,93). Većina sastavnica seruma, koje uzrokuju tzv. učinak matriksa, imaju mali afinitet vezanja za analit ili antitijelo. Obično ta sastavnica maskira analit ili antitijelo, zbog čega izostaje reakcija vezanja analita s antitijelom. Osim svih endogenih elemenata (94), koji uzrokuju inter-individualnu (25) i intra-individualnu varijabilnost (95-97) rezultata imunokemijskih analiza, pojам učinka matriksa mogao bi se proširiti i na egzogene sastavnice, koje se odnose na utjecaj antikoagulansa pri uzorkovanju plazme (98), odnosno na utjecaj aktivatora zgrušavanja krvi i separatora pri uzorkovanju seruma. Heparinska terapija bolesnika s akutnim infarktom miokarda utječe na rezultat određivanja troponina I (96). Heparin se zbog negativnog naboja polianiona, veže s kationima troponina (96). Rezultat toga mogu biti konformacijske promjene molekule troponina ili prikrivanje epitopa koji sudjeluju u imunokemijskoj reakciji s antitijelima reagensa. Osim toga, heparin se različitim afinitetom veže s pojedinim oblicima troponina, koji se mogu naći u krvi bolesnika u različitim fazama nakon infarkta miokarda (95). EDTA može djelovati na oslobađanje slobodnog cTnI iz kompleksa cTnI-troponin C

mor markers (especially CA 19-9, PSA) (34). Tumor marker determination is mostly influenced by possible huge concentration increase (>10,000 fold) that exists in patients with extreme tumor growth. It can occur at the initial laboratory workup of the patient. The cut-off value, where the linearity of turbidimetric methods is lost, is shifted towards higher values by introducing latex particles as antibody carriers. If the reagent manufacturer has not marked the cut-off analyte concentration above which the hook effect occurs, medical biochemist should investigate this and record this data into the manual on work quality of the laboratory. The possibility of the hook effect occurrence is discovered by determining sample with exceptionally high concentration in nondiluted form and in dilutions 1:10 i 1:100 (34,91). If in diluted samples higher values are measured than in nondiluted sample, we are talking about the hook effect. The reliable determination of cut-off concentration follows afterwards. The sample must be diluted until the results of two different dilutions match (taking into consideration the dilution factor). If extreme increased value of the measured analyte is expected, two samples could be prepared – the nondiluted and the diluted one.

The matrix effect

Serum or plasma sample is a complex compound of lipids, proteins, carbohydrates, salt and water. The sum of interferences of all sample components (with exception of analytes), which affect the target analyte to be measured is known as “the matrix effect” (92,93). The most serum components that cause so called matrix effect have low affinity of binding to an analyte or antibody. This component disguises usually the analyte or the antibody causing the absence of the binding reaction of analyte to antibody.

Except of these endogenous elements (94), which cause inter- (25) and intra-individual variability (95-97) of results, the concept of the matrix effect could be widened to exogenous components that relate to the impact of anti-coagulant during plasma sampling (98) or the impact of coagulation activator and separator during serum sampling. Heparin therapy in patients with acute myocardial infarction (AMI) affects the result of determination of troponin I concentration (96). Due to negative charge of polyanions heparin binds with cations of troponin (96). This can result either in conformational changes in troponin molecule or in directly covering epitopes involved in the immunoreaction with antibodies from the reagent. Besides, heparin binds with different affinity to some troponin forms present in patient's blood in different phases after myocardial infarction (95). EDTA can act upon release of free cTnI from calcium ion dependent cTnI-troponin C complex (96) what causes falsely decrease of values

ovisnog o ionima kalcija (96), što uzrokuje lažno smanjene vrijednosti u metodama koje sadrže antitijela usmjerena prema kompleksu troponina. Iako postoje preporuke da se srčani biljezi određuju u plazmi, osobito u hitnoj službi (99), za određivanje troponina uzorak izbora jest serum (može biti uzorkovan s gelom ili bez gela u epruveti) ili uzorak uzorkovan u epruvetu koja sadrži trombin (s ili bez gela) (95).

Gel koji služi kao separator seruma može adsorbirati analit, što može izazvati lažno smanjenu koncentraciju kod određivanja lijekova, primjerice antidepresiva, benzodiazepina (100). Pet do trideset % lijeka može se adsorbirati na gel, a ako uzorak stoji dulje vrijeme (24 h) adsorpcija može iznositi i do 40%. U slučaju da uzorak za analizu treba zamrznuti, serum se mora odvojiti u posebnu epruvetu (101). Činjenica da na tržištu postoji više vrsta epruveta s gelom različite kakvoće, potiče na oprez pri njihovom odabiru, odnosno ukazuje na potrebu njihove validacije pri određivanju pojedinih analita.

Uzrok varijabilnosti rezultata može biti i u matriksu kalibratora (102,103), odnosno kontrolnih uzoraka (20), jer nemaju istovjetan matriks kao biološki uzorak u kojem se neki analit određuju.

Zaključak

Danas se imunokemijske analize ne primjenjuju samo u specijalističkim laboratorijima, nego i u općim medicinsko-biokemijskim laboratorijima (104), a pogotovo u privatnim laboratorijima. Imunometrijskim metodama određuju se hCG (105), hormoni štitnjače (20,106), srčani biljezi (29-31,107), tumorski biljezi (55,73,74,108,109) pa su kod tih analita opisivane i interferencije heterofilnih antitijela. Pojavnost interferencija manja je u analizama koje se primjenjuju dulje vrijeme (proizvođači reagensa i analizatora su ih nastojali otkloniti), nego u analizama koje su kratko u primjeni (110). Osobito treba obratiti pozornost na imunokemijske metode uz krevet bolesnika kod kojih je također opisana interferencija heterofilnih antitijela (111).

Poznavanje brojnih interferencija preduvjet je za njihovo prepoznavanje. Njihovim prepoznavanjem mogu se izbjegći moguće neželjene posljedice važne kako za bolesnika (pogreške u dijagnozi, liječenje i praćenje uspješnosti liječenja, nepotrebna terapija) tako i za zdravstveni sustav (nepotrebna dodatna istraživanja).

Adresa za dopisivanje:

Slavica Dodig
Odjel za kliničko-laboratorijsku dijagnostiku
Dječja bolnica Srebrnjak
Srebrnjak 100
10000 Zagreb
e-pošta: slavica.dodig@zg.t-com.hr

in methods containing antibodies, directed to troponin complex. Despite the recommendations that heart markers should be determined in plasma, especially in emergency department (99), the sample of choice for troponin determination is serum, collected in tubes with or without gel or in thrombin tubes with or without gel (95).

Gel used as serum separator can adsorb analyte what can cause falsely low concentration of drugs prescribing, e.g. antidepressants, benzodiazepine (100). Five to thirty % of the drugs can be adsorbed on gel and if the sample is kept for longer time (24h) adsorption can rise up to 40%. In case the sample for analysis requires freezing, the serum must be placed into a separate test tube (101). There are several different types of test tubes with gel of different quality available on the market today. Therefore, one should be cautious in choosing them and the need for their validation by determination of some analytes is rising.

The cause for result variability may be in the matrix of the calibrator (102,103) or the control samples (17) due to the fact that their matrix is not identical with the biological sample in which some analyte is being determined.

Conclusion

Today, immunoassays are not applied only in specialist laboratories but also in medical biochemical laboratories (104) and especially in private laboratories. Saturation methods are used for determination of hCG (105), thyroid hormones (20,106), cardiac marker (29-31,107), tumor markers (55,73,74,108,109) and in these analytes it has also been reported about interferences of heterophilic antibodies. Prevalence of interference is lower in analyses used for longer period of time (manufacturers of reagents and analysers tended to eliminate them) than in those which are in use for shorter period of time (110). Special attention must be paid to assays with bedside measurements where is also reported about interference of heterophilic antibodies (111).

Knowledge of numerous interferences is a prerequisite for their recognition which helps avoiding possible undesirable consequences important for the patient (diagnostic errors, treatment and monitoring of its efficacy, unnecessary therapy) and for the health care system as well (unnecessary additional researches).

Corresponding author:

Slavica Dodig
Department of Clinical Laboratory Diagnosis
Srebrnjak Children's Hospital
Srebrnjak 100
10000 Zagreb
Croatia
e-mail: slavica.dodig@zg.t-com.hr

Literatura/References

1. Selby C. Interference in immunoassay. *Ann Clin Biochem* 1999;36: 704-21.
2. Ismail AAA, Walker PL, Barth JH, Lewandowski KC, Jones R, Burr WA. Wrong Biochemistry Results: Two Case Reports and Observational Study in 5310 Patients on Potentially Misleading Thyroid-stimulating Hormone and Gonadotropin Immunoassay Results. *Clin Chem* 2002;48:2023-9.
3. Marks V. False-Positive Immunoassay Results: A Multicenter Survey of Erroneous Immunoassay Results from Assays of 74 Analytes in 10 Donors from 66 Laboratories in Seven Countries. *Clin Chem* 2002;48: 2008-16.
4. Preiser W, Brink NS, Hayman A, Waite J, Balfe P, Tedder RS. False-negative HIV antibody test results. *J Med Virol* 2000;60:43-7.
5. Bohner J, von Pape K-W, Hannes W, Stegmann T. False-negative immunoassay results for cardiac troponin I probably due to circulating troponin I autoantibodies. *Clin Chem (letter)* 1996;42:2046.
6. Ismail AA. On the interpretation of affirmative follow-up tests in immunoassays: what must not be done? *Ann Clin Biochem* 2006;43: 249-51.
7. McNeely M. Antibodies: the laboratory depends on them but they can let us down. *Labmedicine* 2002;33:873-6.
8. Cole LA, Rinne KM, Shahabi S, Omrani A. False-positive hCG assay results leading to unnecessary surgery and chemotherapy and needless occurrences of diabetes and coma. *Clin Chem* 1999; 45:313-4.
9. Rotmensch S, Cole S. False diagnosis and needless therapy or presumed malignant diseases in women with false-positive human chorionic gonadotropin concentrations. *Lancet* 2000;355:712-5.
10. Kroll MH, Elin RJ. Interference with clinical laboratory analyses. *Clin Chem* 1994;40:1996-2005.
11. Kricka LJ. Interferences in immunoassays – still a threat. *Clin Chem* 2000; 46:1037-8.
12. Tate J, Ward G. Interferences in Immunoassay. *Clin Biochem Rev* 2004; 25:105-20.
13. Emerson JF, Ngo G, Emerson SS. Screening for interference in immunoassays. *Clin Chem* 2003;49:1163-9.
14. Kricka LJ. Human anti-animal interferences in immunological assays. *Clin Chem* 1999;45:942-56.
15. Ismail AAA. A radical approach is needed to eliminate interferences from endogenous antibodies in immunoassays. *Clin Chem* 2005;51:25-6.
16. Kaplan IV, Levinson SS. When is a heterophile antibody not a heterophile antibody? When it is an antibody against a specific immunogen. *Clin Chem* 1999;45:616-8.
17. Rigo RB, Panyella MG, Bartolomé LR, Ramos PA, Soria PR, Navarro MA. Variations observed for insulin concentrations in an interlaboratory quality program may be due to interferences between reagents and the matrix of the control materials. *Clin Biochem* 2007;40:1088-91.
18. Ylander PJ, Bicskei Z, Hänninen P, Soini JT. Theoretical assessments of errors in rapid immunoassays-how critical is the exact timing and reagent concentrations? *Biophys Chem* 2006;20:141-5.
19. Saidinejad M, Law T, Ewald MB. Interference by carbamazepine and oxcarbazepine with serum- and urine-screening assays for tricyclic antidepressants. *Pediatrics* 2007;120:e504-9.
20. Després N, Grant AM. Antibody interference in thyroid assays: a potential for clinical misinformation. *Clin Chem* 1998;44:440-54.
21. Jones AM, Honour JW. Unusual results from immunoassays and the role of the clinical endocrinologist. *Clin Endocrinol* 2006;64:234-44.
22. Knight AK, Bingeman T, Cole L, Cunningham-Rundles C. Frequent false positive beta human chorionic gonadotropin tests in immunoglobuline A deficiency. *Clin Exp Immunol* 2005;141:333-7.
23. Cavalier E, Carli A, Chapelle JP, Delanaye P. False positive PTH results: an easy strategy to test and detect analytical interferences in routine practice. *Clin Chim Acta* 2008;387:150-2.
24. McCudden CR, Willis MS, Grenache DG. Persistent low concentration of human chorionic gonadotropin in a nonpregnant woman. *Clin Chem* 2008;54:209-14.
25. Berth M, Bosmans E, Everaert J, Dierick J, Schietecatte J, Anckaert E, Delanghe J. Rheumatoid factor interference in the determination of carbohydrate antigen 19-9 (CA19-9). *Clin Chem Lab Med* 2006;44:1137-9.
26. Melanson SE, Lee-Lewandrowski E, Griggs Da, Long WH, Flood JG. Reduced interference by phenothiazines in amphetamine drug of abuse immunoassays. *Arch Pathol Lab Med*. 2006;130:1834-8.
27. Dasgupta A. Endogenous and exogenous digoxin-like immunoreactive substances: impact on therapeutic drug monitoring of digoxin. *Am J Clin Pathol* 2002;118:132-40.
28. Roberts WL, De BK, Coleman JP, Annesley TM. False increased immunoassay measurements of total and unbound phenytoin in critically ill uremic patients receiving fosphenytoin. *Clin Chem* 1999;45:829-37.
29. Panteghini M. Selection of antibodies and epitopes for cardiac troponin immunoassays: Should we revise our evidence-based beliefs? *Clin Chem* 2005;51:803-4.
30. Lum G, Solarz DE, Farney L. False positive cardiac troponin results in patients without acute myocardial infarction. *Labmedicine* 2006;37:546-50.
31. Eriksson S, Halenius H, Pulkki K, Hellman J, Pettersson K. Negative Interference in Cardiac Troponin I Immunoassays by Circulating Tropo-nin Autoantibodies. *Clin Chem* 2005;51:839-47.
32. Duverlie G, Driencourt M, Roussel C, Orfila J. Heterophile IgM, IgA, and IgE antibodies in infectious mononucleosis. *J Allergy Clin Immunol* 2007;119:640-5.
33. Covinsky M, Laterza O, Pfeifer JD, Farkas-Szallasi T, Scott MG. IgM λ antibody to Escherichia coli produces false-positive results in multiple immunoassays. *Clin Chem* 2000;46:1157-61.
34. Wu JT. Quantitative immunoassay: A practical guide for assay establishment, troubleshooting, and clinical application. AAC Press, Washington 2000.
35. Gosling JP. Immunoassays. A practical approach. Oxford: Oxford University Press, 2000.
36. Nilsson C, Seppälä, Pettersson K. Immunological characterization of human luteinizing hormone with special regard to a common genetic variant. *J Endocrinol* 2001;168:107-16.
37. Andrew R. Clinical measurement of steroid metabolism. *Best Practice&Res Clin Endocrinol&Metabolism* 2001;15:1-16.
38. Lee C, Goeger DE. Interference of 6 β-hydroxycortisol in the quantitation of urinary free cortisol by immunoassay and its elimination by solid phase extraction. *Clin Biochem* 1998;31:229-33.
39. Wright JL, Lange PH. Newer potential biomarkers in prostate cancer. *Urology* 2007;9:207-13.
40. Mari A, Iacovacci P, Afferni C, Barletta B, Tinghino R, Di Felice G, et al. Specific IgE to cross-reactive carbohydrate determinants strongly affect the in vitro diagnosis of allergic diseases. *J Allergy Clin Immunol* 1999;103:1005-11.
41. Cantisani A, Giuffrida, Fabris C, Bertino E, Coscia A, Oggero R, et al. Detection of specific IgE to human milk proteins in sera of atopic infants. *FEBS Lett* 1997;412:515-7.
42. Aalberse RC. Allergens from mites: implications of cross-reactivity between invertebrate antigens. *Allergy* 1998;53:47-8.
43. Chou MC, Yuo CY, Lan WS, Huang SP. Serial IgE of atopic individuals exhibit the reactivity to multiple allergens. *J Biomed Lab Sci*. 2001;13:1-5.
44. Fuchs T, Spitzauer S, Vente C, Hevier J, Ksziotis S, Rumpold H, et al. Natural latex, grass pollen, and weed pollen share IgE epitopes. *J Allergy Clin Immunol* 1997;100:356-64.
45. Adéodoin J, Johansson SGO, Grönlund H, van Hage M. Interference in immunoassays by human IgM with specificity for the carbohydrate moiety of animal proteins. *J Immunol Methods* 2006;310:117-25.
46. Hemmer W, Focke M, Kolarich D, Wilson IB, Altmann F, Wohrl S, et al. Antibody binding to venom carbohydrates is a frequent cause for

- double positivity to honeybee and yellow jacket venom in patients with stinging-insect allergy. *J Allergy Clin Immunol* 2001;108:1045-52.
47. Ito K, Morishita M, Ohshima M, Sakamoto T, Tanaka A. Cross-reactive carbohydrate determinant contributes to the false positive IgE antibody to peanut. *Allergol International* 2005;54:387-92.
 48. Van Ree A, Aalberse RC. Specific IgE without clinical allergy. *J Allergy Clin Immunol* 1999;103:1000-1.
 49. Steimer W. Performance and specificity of monoclonal immunoassays for cyclosporine monitoring: how specific is specific? *Clin Chem* 1999;45:371-81.
 50. Lichteblawner MR, Mencken T, Tully R, Petosa M. False-positive immunochemical screen for methadone attributable to metabolites of verapamil. *Clin Chem* 1998;44:1039-41.
 51. Daher R, Haidar JH, Al-Amin H. Rifampin Interference with Opiate Immunoassays. *Clin Chem* 2002;48:203-4.
 52. Valdes R Jr, Jortani SA. Unexpected suppression of immunoassay results by cross reactivity; now a demonstrated cause for concern. *Clin Chem* 2002;48:405-6.
 53. Ellis MJ, Livesey JH. Techniques for identifying heterophile antibody interference are assay specific: study of seven analytes on two automated immunoassay analyzers. *Clin Chem* 2005; 51: 639-41.
 54. Choy EH, Schantz A, Pitzalis C, Kingsley GH, Panayi GS. The pharmacokinetics and human anti-mouse antibody response in rheumatoid arthritis patients treated with a chimeric anti-CD4 monoclonal antibody. *Rheumatology* 1998;37:801-2.
 55. Reinsberg J, Nocke W. Falsely low results in CA 125 determination due to anti-idiotypic antibodies induced by infusion of [131I]Fab'2 fragments of the OC125 antibody. *Eur J Clin Chem Clin Biochem* 1993;31:323-7.
 56. McMillin A, Owen WE, Lambert TL, De BK, Frank EL, Bach PR, et al. Comparable Effects of DIGIBIND and DigiFab in Thirteen Digoxin Immunoassays. *Clinical Chemistry* 2002;48:1580-4.
 57. Dasgupta A, Reyes MA. Effect of Brazilian, Indian, Siberian, Asian, and North American Ginseng on Serum Digoxin Measurement by Immunoassays and Binding of Digoxin-like Immunoreactive Components of Ginseng With Fab Fragment. *Am J Clin Pathol* 2005;124:229-36.
 58. Ward G, McKinnon L, Badrick T, Hickman PE. Heterophilic antibodies remain a problem for the immunoassay laboratory. *Am J Clin Pathol* 1997;108:417-21.
 59. Nahm MH, Hoffman JW. Heterophile antibody: phantom of the immunoassay. *Clin Chem* 1990;36:829.
 60. Levinson SS, Miller JJ. Towards a better understanding of heterophile (and the like) antibody interference with modern immunoassays. *Clin Chim Acta* 2002;325:1-15.
 61. Kuroki M, Matsumoto Y, Arakawa F, Haruno M, Murakami M, Kuwahara M, et al. Reducing interference from heterophilic antibodies in a two-site immunoassay for carcinoembryonic antigen (CEA) by using a human/mouse chimeric antibody to CEA as the tracer. *J Immunol Methods* 1995;180:81-91.
 62. Higgins JP, Higgins JA. Elevation of cardiac troponin I indicates more than myocardial ischemia. *Clin Invest Med*. 2003;26:133-47.
 63. Jeremias A, Gibson CM. Narrative review: Alternative causes for elevated cardiac troponin levels when acute coronary syndromes are excluded. *Ann Int Med* 2005;142:786-92.
 64. Preissner CM, O'Kane DJ, Singh RJ, Morris JC, Grebe SK. Phantoms in the assay tube: heterophile antibody interferences in serum thyroglobulin assays. *J Clin Endocrinol Metab* 2003;88:3069-74.
 65. Giovannella L, Ghelfo A. Undetectable serum thyroglobulin due to negative interference of heterophile antibodies in relapsing thyroid carcinoma. *Clin Chem* 2007;53:1871-2.
 66. Luzzi VI, Scott MG, Gronowski AM. Negative thyrotropin assay interference associated with an IgG kappa paraprotein. *Clin Chem* 2003;49:709-10.
 67. Eriksson S, Junikka M, Laitinen P, Majamaa-Voltti K, Alfthan H, Petersson K. Negative interference in cardiac troponin I immunoassays from a frequently occurring serum and plasma component. *Clin Chem* 2003;49:1095-104.
 68. Martins TB, Pasi BM, Litwin CM, Hill HR. Heterophile antibody interference in a multiplexed fluorescent microsphere immunoassay for quantitation of cytokines in human serum. *Clin Diagn Lab Immunol* 2004;11:325-9.
 69. Kellar KL, Kalwar RR, Dubois KA. Multiplexed fluorescent bead-based immunoassays for quantitation of human cytokines in serum and culture supernatants. *Cytometry* 2001;45:27-36.
 70. Banks RE, Measurement of cytokines in clinical samples using immunoassays: problems and pitfalls. *Crit Rev Clin Lab Sci* 2000; 37:131-82.
 71. Halsall DJ, Mangi M, Soos M, Fahie-Wilson MN, Wark G, Mainwaring-Burton R, et al. Hypoglycemia due to an insulin binding antibody in a patient with an IgA- κ myeloma. *J Clin Endocrinol Metab* 2007;92:2013-6.
 72. Zouwail SA, O'Toole AM, Clark PMS, Begley JP. Influence of Thyroid Hormone Autoantibodies on 7 Thyroid Hormone Assays. *Clin Chem* 2008;54:927-8.
 73. Oei AI, Sweep FC, Massuger LF, Olthaar AJ, Thomas CM. Transient human anti-mouse antibodies (HAMA) interference in CA 125 measurements during monitoring of ovarian cancer patients treated with murine monoclonal antibody. *Gynecol Oncol* 2008;109:199-202.
 74. Bertholf RL, Johannsen L, Benrubi G. False elevation of serum CA-125 level caused by human anti-mouse antibodies. *Ann Clin Lab Sci* 2002;32:414-8.
 75. Oldham RK. Monoclonal Antibodies in Cancer Therapy: 25 Years of Progress. *J Clin Oncol* 2008;26:1774-7.
 76. Dodig S, Richter D, Ćepelak I, Benko B. Anti-IgE therapy with omalizumab in asthma and allergic rhinitis. *Acta Pharm* 2005;55:123-38.
 77. Schulze-Koops H, Lipsky PE. Anti-CD4 Monoclonal Antibody Therapy in Human Autoimmune Diseases. In: Fathman CG. (ed): *Biologic and Gene Therapy of Autoimmune Disease*. Curr Dir Autoimmun. Basel, Karger, 2000, vol 2, pp 24-49.
 78. Reinsberg J. Different efficacy of various blocking reagents to eliminate interferences by human antimouse antibodies with a two-site immunoassay. *Clin Biochem* 1996;29:145-8.
 79. Butler SA, Cole LA. Use of heterophilic antibody blocking agent (HBT) in reducing false-positive hCG results. *Clin Chem* 2001;47:1332-3.
 80. Bjerner J, Bormer OP, Nustad K. The War on Heterophilic Antibody Interference. *Clin Chem* 2005;51:9-11.
 81. Warren J, Bjerner J, Paus E, Bormer OP, Nustad K. Use of an *in vivo* biotinylated single-chain antibody as capture reagent in an immunometric assay to decrease the incidence of interference from heterophilic antibodies. *Clin Chem* 2005;51:830-8.
 82. Wu JT, Christensen SE. Effect of different test designs of immunoassays on "hook effect" of CA 19-9 measurement. *J Clin Lab Anal* 1991;5: 228-32.
 83. Unnikrishnan AG, Rajaratnam S, Seshadri MS, Kanagasabapathy AS, Stephen DC. The 'hook effect' on serum prolactin estimation in a patient with macroadenoma. *Neurol India* 2001;49:78-80.
 84. Amarasing FS, Wilson GS. Studies of the 'hook' effect in the one-step sandwich immunoassay. *J Immunol Methods* 1992;151:47-66.
 85. Levavi H, Neri A, Bar J, Regev D, Nordenberg J, Ovadia J. "Hook effect" in complete hydatidiform molar pregnancy: a falsely low level of β -HCG. *Obstet Gynecol* 1993;82:720-1.
 86. Van Lente F. Light scattering immunoassays. Rose NR de Macario EC Folds JD Lane HC Nakamura RM eds. *Manual of clinical laboratory immunology*, 5th ed 1997:13-19 ASM Press Washington.
 87. Jury DR, Mikkelsen DJ, Dunn PJ. Prozone effect and the turbidimetric measurement of albumin in urine. *Clin Chem* 1990;36:1518-9.
 88. Killeen AA, Ramey ML, Dean JJ. High-dose hook effect in an immunoluminometric thyrotropin assay: the open-faced sandwich artefact. *Ann Clin Biochem* 1993;30:413-4.
 89. Charrie A, Charriere G, Guerrier A. Hook effect in immunometric assays for prostate-specific antigen. *Clin Chem* 1995;41:480-1.
 90. Fernando SA, Wilson GS. Studies of the 'hook' effect in the one-step sandwich immunoassay. *J Immunol Methods* 1992;151:47-66.

91. Cole TG, Johnson D, Eveland BJ, Nahm MH. Cost-effective method for detection of "hook effect" in tumor marker immunoassay. *Clin Chem* 1993;39:695-6.
92. Wood GW. "Matrix effect" in immunoassay. *Scand J Clin Lab Invest* 1991;51:105-12.
93. Yoshida H, Imafuku Y, Nagai T. Matrix effects in clinical immunoassays and the effect of preheating and cooling analytical samples. *Clin Chem Lab Med* 2004;42:51-6.
94. Feng-Bo Wu, You-Feng He, Shi-Quan Han. Matrix interference in serum total thyroxin (T4) time-resolved fluorescence immunoassay (TR-FIA) and its elimination with the use of streptavidin-biotin separation technique. *Clin Chim Acta* 2001;308:117-26.
95. Gerhardt W, Nordin G, Herbert AK, Burzell BL, Isaksson A, Gustavsson E, et al. Troponin T and I assays show decreased concentrations in heparin plasma compared with serum: lower recoveries in early than in late phases of myocardial injury. *Clin Chem* 2000;46:817-21.
96. Katrukha AG, Bereznikova AV, Esakova TV, Petterson K, Lövgen T, Severina ME, et al. Troponin I is released in bloodstream of patients with myocardial infarction not in free form but as a complex. *Clin Chem* 1997;43:1379-85.
97. Hirsh J, Raschke R, Warkentin TE, Dalen JE, Deykin D, Foller L. Heparin: mechanism of action, pharmacokinetics, dosing, considerations, monitoring, efficacy, and safety. *Chest* 1995;108(Suppl 4):258-75.
98. Wu AHB, Fen YJ, Moore R, Apple FS, McPherson PH, Buechler KF, Bodor G. Characterization of cardiac troponin subunit release into serum after acute myocardial infarction and comparison of assays for troponin T and I. *Clin Chem* 1998;44:1198-208.
99. Wu HB, Apple FS, Gibler BW, Jesse RL, Warshaw MM, Valdes R. National Academy of Clinical Biochemistry Standards of Laboratory Practice: recommendations for the use of cardiac markers in coronary artery diseases. *Clin Chem* 1999;45:1104-21.
100. Karppi J, Akerman KK, Parviainen M. Suitability of collection tubes with separator gels for collecting and storing blood samples for therapeutic drug monitoring (TDM). *Clin Chem Lab Med* 2000;38:313-20.
101. Rosa-Fraile M, Sampedro A, Rodriguez-Granger J, Camacho E, Manrique E. Suitability of Frozen Serum Stored in Gel Separator Primary Sampling Tubes for Serological Testing. *Clin Diagn Lab Immunol* 2004;11:219-21.
102. Rej R, Drake P. The nature of calibrators in immunoassays: are they commutable with test samples? Must they be? *Scand J Clin Lab Invest Suppl* 1991;205:47-54.
103. Gaines-Das RE, Brettschneider H, Bristow AF. The effects of common matrices for assay standards on performance of 'ultra sensitive' immunoassays for TSH. Report of a joint WHO/IFCC collaborative study. *Clin Chim Acta* 1991;16:203:5-15.
104. Hubl W, Zogbaum M, BOyd JC, Savory J, Schubert M, Meyer D, Dermant T. Evaluation of analytical methods and workflow performance of the Architect ci8200 integrated serum/plasma analyzer system. *Clin Chim Acta* 2005;357:43-54.
105. Esfandiari N, Goldberg JM. Heterophile antibody blocking agent to confirm false positive serum human chorionic gonadotropin assay. *Obstet Gynecol* 2003;101:1144-6.
106. Norden AG, Jackson RA, Norden LE, Griffin AJ, Barnes MA, Little JA. Misleading results from immunoassays of serum free thyroxine in the presence of rheumatoid factor. *Clin. Chem* 1997; 43: 957-62.
107. Fitzmaurice TF, Brown C, Rifai N, Wu AHB, Yeo K-TJ. False increase of cardiac troponin I with heterophilic antibodies. *Clin Chem* 1998;44:2212-4.
108. Morrissey NE, Quadri SF, Kinders R, Brigham C, Rose S, Blend MJ. Modified method for determining carcinoembryonic antigen in the presence of human anti-murine antibodies. *Clin Chem* 1993;39:522-9.
109. Preissner CM, Dodge LA, O'Kane DJ, Singh RJ, Grebe SKG. Prevalence of heterophilic antibody interference in eight automated tumor marker immunoassays. *Clin Chem* 2005; 51:208-10.
110. Ismail AA, Walker PL, Cawood ML, Barth JH; Interference in immunoassay is an underestimated problem. *Ann Clin Biochem* 2002;39:366-73.
111. White GH, Tideman PA. Heterophilic antibody interference with CARDIAC T quantitative rapid assay. *Clin Chem* 2002;48:201-3.