

The perfect osseointegration process of the dental implants depends among other factors on the surface characteristics of the titanium. In this study enlarged mechanical roughness was produced by a laser-based technique, in order to decrease the healing period of the implant.

There are different ways of forming laser induced surface structures. In the case of mask projection techniques the surface can be modified in larger areas and surface patterns. An ArF nanosecond excimer laser was used in the experiments because of the advantageous properties of the excimer beams. Effective polishing by homogeneous laser illumination in the 3-5 J/cm<sup>2</sup> fluence range was performed as confirmed by SEM and AFM studies. Holes of about 20 mm diameter and 10 mm in depth, with high aspect ratio and protrusions around the edges were ablated into the titanium surface with subsequent pulses of ns ArF excimer laser. To avoid easily breakable protrusions we applied excimer pulse durations of 0.5 picoseconds. In this case we obtained melting- and ridge-free ablation of titanium. The laser treatment influenced the chemical composition of the surface in two respects. On the one hand it removed carbonaceous contamination as indicated by XPS and XRD measurements, demonstrating that cleaning of the surface does not alter the original crystalline structure. On the other hand, XPS measurements proved that pulsed laser oxidation in air increased the thickness of the surface oxide layer, promoting better osseointegration.

## 40. Porcelain-Fused-to-Titanium Restorations From Implant Level - - Preliminary Observations

**Stendera P, Matenko D, Grochowski P.**  
*"Dental Service" Clinic, Warsaw, Poland*

Until a few years ago use of the abutment was necessary for making fixed restorations retained on Brånemark implants. For multiple-unit restorations, abutments served as a "misfit eliminating zone" for the framework cast with gold alloy. Single crowns were made as restorations cemented on prefabricated abutments. Thanks to the use of modern laboratory methods, it is now possible to produce multiple-unit restorations retained directly on implants with the metal part made of titanium alloy. One such method is the All-in-One system, in which the metal part of the restoration is made of one piece of titanium, with the use of a computer aided milling machine. Thanks to the development in porcelain fusing to titanium, it is also possible to make single crowns as a screw-retained

restoration seated directly on an implant. For this purpose, TiAdapt abutments were used with own modification of UCLA-type crown production.

The paper describes some advantages and disadvantages of the above mentioned systems and basics of their use.

The study group consisted of 22 restorations (12 bridges and 10 crowns) made for 14 patients. Most of them (18) were made for the lower jaw. All bridges were two- or three-unit restorations. The observation period was from three months up to two years. Complications were observed in two cases. Both of them consisted of cracks in the porcelain. Due to the design of the restorations, the complications were eliminated without any problems. A possible cause of the porcelain cracking was the excessive bulk of the porcelain in areas where there had been lack of metal support for the veneering layer.

Due to the easy application and positive aesthetic and functional results the described methods seem to be worth recommending, with some clinical limitations.

## 41. Three-Year Clinical Evaluation of in Ceram Zirconia Bridges.

**Suárez MJ, Lozano JF, Salido MP, Pradíes G,  
Martínez F.**  
*Department of Prosthodontics, Faculty of Dentistry,  
Complutense University, Madrid, Spain*

**INTRODUCTION:** The demand for esthetic restorations has led to the development of materials that are metal free. These materials must have adequate strength to be an alternative for the fabrication of fixed partial dentures in posterior segments of the dental arches.

**PURPOSE:** The aim of the present study was to investigate in a long-term perspective whether the strength of the In Ceram Zirconia material is sufficient for use in posterior bridges.

**MATERIAL AND METHODS:** Sixteen patients, who were between 23 and 50 years of age, with indications for a fixed denture replacing premolar or molar, were examined for participation in the study. Eighteen bridges were constructed with one or two pontics and two abutments, one on each side of the pontic. The patients were informed about risks of, and alternatives to, the proposed therapy. The supporting teeth were prepared with chamfer finish line and lack of sharp line angles. Impressions were made with a rigid standard tray with an A-silicone putty soft and light-body materials (Aquasil, Dentsply). The laboratory procedures were performed by a laboratory autho-