## **Review**

# Blood alcohol concentration in the clinical laboratory: a narrative review of the preanalytical phase in diagnostic and forensic testing

#### Cristiano lalongo\*

Department of Experimental Medicine, Policlinico Umberto I, 'Sapienza' University, Rome, Italy

\*Corresponding Author: cristiano.ialongo@uniroma1.it

#### Abstract

The analysis of blood alcohol concentration (BAC), a pivotal toxicological test, concerns acute alcohol intoxication (AAI) and driving under the influence (DUI). As such, BAC presents an organizational challenge for clinical laboratories, with unique complexities due to the need for forensic defensibility as part of the diagnostic process. Unfortunately, a significant number of scientific investigations dealing with the subject present discrepancies that make it difficult to identify optimal practices in sample collection, transportation, handling, and preparation. This review provides a systematic analysis of the preanalytical phase of BAC that aims to identify and explain the chemical, physiological, and pharmacological mechanisms underlying controllable operational factors. Nevertheless, it seeks evidence for the necessity to separate preanalytical processes for diagnostic and forensic BAC testing. In this regard, the main finding of this review is that no literature evidence supports the necessity to differentiate preanalytical procedures for AAI and DUI, except for the traceability throughout the chain of custody. In fact, adhering to correct preanalytical procedures provided by official bodies such as European federation of clinical chemistry and laboratory medicine for routine phlebotomy ensures both diagnostic accuracy and forensic defensibility of BAC. This is shown to depend on the capability of modern pre-evacuated sterile collection tubes to control major factors influencing BAC, namely non-enzymatic oxidation and microbial contamination. While certain restrictions become obsolete with such devices, as the use of sodium fluoride (NaF) for specific preservation of forensic BAC, this review reinforces the recommendation to use non-alcoholic disinfectants as a means to achieve "error-proof" procedures in challenging operational environments like the emergency department.

Keywords: preanalytical phase; blood alcohol concentration; specimen handling; humans; substance abuse detection

Submitted: July 18, 2023

Accepted: August 20, 2023

#### Introduction

Ethyl alcohol ( $C_2H_5OH$ , CAS 64-17-5) or ethanol (EtOH) or just "alcohol", is a liquid organic compound at room temperature that forms the basis of various aqueous beverages (spirits, wine, or beer) consumed by half of the world's population over the age of 15, the majority of whom reside in the Americas, Europe, and the Western Pacific (1). For individuals up to the age of 49, the consumption of EtOH is the leading risk factor for premature death and disability, with mortality rates even exceeding those of diabetes, tuberculosis, and human immunodeficiency virus (HIV)/acquired immunodeficiency syndrome (AIDS) (2). Blood alcohol concentration (BAC) analysis is a toxicological test associated with two important conditions: acute alcohol intoxication (AAI) and driving under the influence (DUI). Acute alcohol intoxication is a clinical emergency that affects 7.5% of drinkers and causes 2200 deaths *per* year in countries such as the United States (1,3). The diagnostic determination of BAC is routinely performed by means of enzymatic reaction quantified *via* automated spectrophotometry (*i.e.*, clinical chemistry auto-analysers), using serum or plasma as the elective matrix being thus referred to as serum alcohol concentration (SAC) or plasma alcohol concentra-

https://doi.org/10.11613/BM.2024.010501

Biochem Med (Zagreb) 2024;34(1):010501

<sup>©</sup>Copyright by Croatian Society of Medical Biochemistry and Laboratory Medicine. This is an Open Access article distributed under the terms of the Creative Commons Attribution License (https://creativecommons.org/license/by/.40/) which permits users to read, download, copy, distribute, print, search, or link to the full texts of these articles in any medium or format and to remix, transform and build upon | the material, provided the original work is properly cited and any changes properly indicated.

tion (PAC). Driving under the influence, on the other hand, is a condition that, although not necessarily associated with fatal intoxication, results in cognitive impairment responsible for 25% of fatal road accidents, causing approximately 10,000 deaths *per* year in both Europe and the United States (2-5). The forensic determination of BAC is carried out by means of gas-chromatographic separation after headspace extraction (HS-GC) with either mass spectrometric (MSD) or flame ionization (FID) detection, using whole blood as the matrix of choice.

Operationally, beside AAI, cases a hospital and its clinical laboratory can be involved at various levels in the analysis of DUI, being in charge of sample collection alone or of the full confirmatory analysis. However, while the forensic analytical phase may be outsourced due to the need for specialized instrumentation and expertise, the resources of the preanalytical phase - from the personnel responsible for sample collection to the refrigerators used for sample storage - are often shared between diagnostic and forensic BAC testing. This has led to extensive discussions among laboratory specialists, which can be summarized as the need to identify, on one hand, the forensic defensibility characteristics of data produced by the diagnostic process and, on the other hand, the sustainability in a clinical environment of a specific process that guarantees the legal validity of the results (6-11).

To gain control over the preanalytical phase through operational procedures, it is necessary to understand the role and significance of various factors in determining BAC. Considering that the preanalytical phase in BAC accounts for 40% of the analytical result, compared to 20% for biological variability, one must have a comprehensive view of which factors are actually controllable (12). Notable contributions in this regard are provided by the scientific literature on forensic subjects (13). However, specifically for the clinical laboratory scientist, there is only one official document issued by the Clinical and laboratory standards institute (CLSI) (14). However, this document suffers from a series of limitations: it has never been revised since its initial publication in 1997, it is not currently officially available, and, most importantly, it only covers the preanalytical phase in three paragraphs (2.3 - 2.5) based on just six references.

The purpose of this literature review is therefore to provide the laboratory professional with the most comprehensive understanding of the factors determining the preanalytical phase of BAC, especially considering the coexistence of diagnostic and forensic processes in the clinical setting. A series of appendices to the text provides further insights into remarkable topics related with the BAC and the investigation thereof.

## Literature search

Pubmed, Google Scholar and MEDLINE were searched for papers published until December 2022 with no restriction on language (see Appendix A). The search strategy based on a categorization of the subject according to a suitable model of the pre-analytics of a drug testing based on four major topics (namely, sampling, handling, contamination and matrix) is also represented in Figure 1. A typical query used for a preliminary search was as follows: ("blood alcohol" OR "blood ethanol") AND "sampl\*" AND ("stabil\*" OR "stor\*" OR "temperature" OR "contaminat\*" OR "factor\*"). The search was then refined within each topic adjusting by the use of more selective terms (e.g., "stabiliz\*", "preserv\*", "additive", "sealing", "leak\*" for the handling factor or "haemol\*", "clot\*", "lipem\*", "icter\* for the matrix factor"). The literature search was further extended reviewing bibliography within each article issued before 1980s in order to retrieve any cited source that was eventually not electronically indexed (see Appendix A). Studies concerning *post-mortem* specimens or animals were excluded.

## **Pre-analytical factors**

#### Sampling time

Ethanol is a very low molecular weight compound (46.07 g/mol and density of 0.789 g/mL at 20 °C) that can freely diffuse through the cell membrane (13). When ingested as a diluted water solution, it is absorbed *per* passive diffusion, first in the stom-

2



**FIGURE 1.** Fish-bone diagram of the major preanalytical factors of blood alcohol concentration (BAC) testing. To each factor (lower-case bold) is associated a source of bias (upper-case italics) with the sign indicated in parenthesis (the symbol " $\pm$ " means that the exact sign depends on conditions as explained in the text).

ach and then extensively in the intestine (13). The peak of blood absorption depends on the timing of gastric emptying but not on the concentration or the volume of the consumed beverage due to the very large absorptive surface of the proximal small bowel (15,16). Accordingly, the time to reach body fluids equilibrium is erratic and depends on both the stomach emptying and the variations of the blood flow of the gut (17).

In general, the fasting peak blood is observed within 45-60 minutes after finished drinking, however in some subjects it may already be present by as early as 15 minutes or delayed to up to 120 minutes from ingestion (13). In this phase, where EtOH is equilibrating between blood and tissues, there is a negative veno-arterous bias (- $\Delta_{V/A}$ ) up to - 0.20 g/L in the same tributary area (*e.g.*, cephalic vein *vs*. radial artery), and a positive veno-venous bias  $(+\Delta_{V/V})$  between proximal and distal tributary areas (*e.g.*, up to 0.06 g/L forearm *vs*. opposite foot) (17-20). Venous BAC is also affected by large erratic fluctuations caused by the equilibration phenomena (21).

In the post-absorption phase that begins about 90 minutes post-drinking, the veno-arterous bias changes the sign  $(+\Delta_{A/V})$  and no veno-venous bias  $(\Delta_{V/V} \approx 0)$  is usually observed (17-20). In this phase, the venous BAC reflects the distribution equilibrium between blood and tissues, as in the brain where it exerts its psychoactive action. In this phase, if blood is recollected within 10 minutes from the previous sample, the resampling bias lays

within the physiological random fluctuations (averaging  $\pm$  0.008 g/L and  $\pm$  0.010 g/L for venous and arterial blood, respectively) (13,22).

It should be noted that due to the kinetics of distribution to and from the tissues, the capillary blood shows an additive negative bias ranging by - 0.04 to - 0.06 g/L (17,23).

### Sampling site

The superficial veins laying in the antecubital fossa of the forearm represent the elective sites of routine phlebotomy, however they are characterized by remarkable topographical variability (both within- and between-subject) and more anastomoses than the deeper arteries (24). Thus, even when  $\Delta_{V/V} \approx 0$ , it can be observed with fairly large prevalence (almost 50%) an inter-arm bias ( $\Delta_{I/R}$ ) that ranges between 0.008 g/L and 0.024 g/L with significant between-subject variability (coefficient of variation, CV≈60%) (12,17,25). Despite the size of this bias is almost the same before and after the absorption peak, its fluctuations are much larger in the absorption phase (17). Remarkably, this bias has unpredictable direction as it is unrelated with handedness, gender, pattern of the superficial veins, level of the blood gasses and haematocrit (12,17,25).

### Sampling technique

Since water content of blood and tissues shows very small within-subject variability (CV < 3.0%), any mechanical factor disturbing the perfusion flow can affect the equilibration of EtOH and in turn the BAC (22,26,27). For instance, a difficult positioning of an indwelling sampling device (catheter, butterfly needle) can increase up to five-fold the time-independent random BAC differences observed in the post-absorption phase (22).

The vasoconstriction of the tributary area induced by cooling (*i.e.*, at 13-15 °C) doubles both the size and the duration of the  $\Delta_{A/V}$  when the EtOH is administered by infusion, whereas the vasodilation (*i.e.*, warming at 60 °C) halves only the size of the  $\Delta_{A/V}$  (28). However, when EtOH is ingested, the duration (1 to 5 minutes) and the pressure applied to induce the venous stasis (60 to 100 mmHg) as well as the local ice-cooling of the skin do not have an effect over the BAC, while the body positioning (orthostatic reaction) and the physical activity produce a negative bias (29). This apparent contradiction between infused and ingested EtOH may eventually depend on the fact that the post-absorption route of distribution plays an unaddressed experimental and physiological role.

### **Chemical contamination**

Ethanol volatizes quickly from surfaces (the wellknown "cooling effect" experienced after skin cleansing) with half-life on skin of 11.7 second, so that theoretically is necessary to expose 1000 cm<sup>2</sup> of skin to 70% v/v alcoholic solution to achieve a BAC of 0.06 g/L (30).

Experimentally, a chemical contamination during sampling requires that the needle is suctioning while it is in direct contact with the alcoholic antiseptic (e.g., pressing the soaked swab to stop bleeding while withdrawing the needle) (31-33). Otherwise, (e.g., without pressing the soaked swab on the site of phlebotomy), with pre-evacuated tubes there is no contamination when the excess antiseptic is used (i.e., 2 mL) and alcohol is not allowed to dry off (i.e., 5 seconds waiting) (34). Likewise, the use of a syringe and thus of a controlled suction makes actually difficult to contaminate deliberately the specimens even pouring the alcohol directly onto the skin and inserting the needle shortly afterwards (35). Hence, if the correct amount of antiseptic is used (i.e., 1 mL) and it is allowed to dry off (i.e., 1 minute), contamination has only 5% probability or less to happen (36,37). However, the spurious BAC due to a chemical contamination is erratic and unpredictable (e.g., depending on the degree of swab squeezing), and values ranging from 0.005 g/L to up to 6.0 g/L have been reported (33,35,36).

It must be noted that when the experimentation involves inebriated subjects, the EtOH swabbing seems to give an average increase of 0.05 g/L of the BAC regardless of whether a pre-evacuated tube or a syringe was used (38-40). As already discussed for the mode of blood sampling, the discrepancy that occurs when the study involves inebriated subjects may rather be the effect of some uncontrolled experimental factor.

The effect of the chemical contamination depends on the analytical method used for measuring the BAC when skin disinfection is made with alcohol like isopropyl or amyl, since the spectrophotometric enzymatic assay based on yeast alcohol dehydrogenase (ADH) is highly selective for EtOH (37,41-43). Indeed, contaminating with isopropanol (*i.e.*, 2-propanol) gives a negative bias when the BAC quantitation *via* the HS-GC uses this alcohol in place of the isomer n-propanol (1-propanol) as the internal standard (44).

#### **Microbial contamination**

Even though the pre-evacuated tube used for blood sampling is a closed and sterile system that avoids environmental contamination during and after the blood withdrawal, the incorrect skin disinfection can be the source of microbes' contamination during phlebotomy. Indeed, *Proteobacteria*, *Bacterotides* and *Staphylococcaceae* colonize the moisty skin of the antecubital fossa (45,46). Among the fungi, *Malassezia* predominates on *Candida* (47). All such microbes are able to synthesize anaerobically the EtOH *via* the fermentation (with production of small amounts of ethyl acetate as by-product), except for *Malassezia* that can only hydrolyse the fatty acids ethyl esters (48,49).

In a freshly collected blood sample, the contaminating microbes are in the disadvantage respect to the far more numerous erythrocytes for the uptake of glucose. Therefore, in such unfavourable conditions of substrates availability, the microbial ADH operates for the salvage pathway by reducing EtOH to restore nicotinamide adenine dinucleotide NAD<sup>+</sup> from NADH + H<sup>+</sup> (50-52). As the consequence, the bias of the BAC from a contaminated specimen is expected to be negative, except for the case of supplementation of the sample with extra-glucose as in a banked blood bag (53,54). Accordingly, the BAC of an EtOH-free (or below the detection limit) specimen remains unchanged regardless such factor as the use of sodium fluoride (NaF) preservative, the duration and temperature

of the storage and the conditions of handling (*e.g.*, heat exposure, repeated uncapping and sampling) (55-60). Likewise, the environmental contamination of the blood matrix during the sampling and processing operations (*e.g.*, uncapping, decanting, pipetting) is unlikely to occur in a typical laboratory setting even when the sample is forcedly exposed or handled carelessly (57).

The NaF represents the major anti-microbial agent used for stabilizing the BAC and it has been recommended since very early in the scientific literature (61). At a concentration ranging within 200 mM and 300 mM ( $\approx$ 1% w/v), it causes the cell death within 24 hours, while at lower concentrations (e.a., 100 mM) it can take up to 72 hours (62). Hence, at least within 24 hours the preservation of BAC is NaF-independent (51). Nonetheless, such microbes as Pseudomonas and Serratia can survive to NaF exposure up to 600 mM ( $\approx 2\%$  w/v) (52). Hence, NaF acts as a second-line defense against contamination but is not superior to the primary sterility of the collection tube. Remarkably, in an ordinary tube used for glucose testing, the concentration of NaF is 60 mM (i.e., 0.25% w/v or 2.5 mg/mL whole blood) and thus it is inadequate for preventing the growth of the microbes.

## Sample container

Ethanol is an organic compound whose octanol/ water partition coefficient (logP<sub>o/w</sub>) lays close to -0.3, therefore it cannot be adsorbed into the gel separator of clinical chemistry tubes as it requires  $logP_{o/w} > 3$  (63-66). Accordingly, no adsorption bias is observed when the serum BAC is measured in plain or gel separator tubes (67,68). On the contrary, as a volatile organic solvent, the EtOH permeates the walls of tube made of single-layered polypropylene, but not the inner-layered polyethylene terephthalate and the glass borosilicate (64,69-71). In general, the permeation bias between a double-walled plastic and a glass tube rests within 1% of BAC regardless of the storage temperature (69,72). This condition is also unaffected by the extension of the contact surface of blood on the tube walls (i.e., whether the tube is stored standing up or laying down) (70).

Occasionally, the gel separator can be the source of some chemical contamination, especially in the past due to the solvents used for the manufacturing of the rubber stoppers (unidentified substance) and gel separator of serum tubes (toluene, 1-butanol, ethylbenzene and xylene) (73-75). This was also found in whole blood sodium fluoride/oxalate (NaF/Ox) tubes containing traces of isobutylene (76). However, none of the above compound interferes with the spectrophotometric enzymatic assay of BAC. Conversely, chemical contaminants released from gel separator tubes and able to interfere with the HS-GC analysis carried out without the mass selective detector (MSD) were also reported recently. Particularly, these were a 1-propanol-like substance and ethyl chloride (chloroethane,  $C_2H_5Cl$ ), both capable to positively bias the BAC as almost co-eluting with EtOH (77-79).

#### Dispersion

The complete solubility of EtOH into water is due to the presence of the hydroxyl moiety on the very short carbon chain, so that the water/air partition coefficient ( $logP_{w/a}$ ) measured in distilled water does not significantly differ from whole blood and plasma ( $logP_{w/a} \approx 3.33$ ,  $logP_{w/a} \approx 3.24$  and  $logP_{w/}$  $_a \approx 3.31$ , respectively) (80,81). Consequently, the evaporation bias of an uncapped serum or heparinized plasma tube resting at 22.1 - 25.1 °C temperature and 55% maximum air humidity averages - 3.0% within 30 minutes, - 5.0% after 1 hour and reaches a maximum of - 10.0% after 3 hours (82).

In a capped tube, EtOH collects into the air volume above the specimen (so-called "headspace"), reaching a concentration (*i.e.*, partial pressure) that only depends on the temperature as explained by the Henry-Dalton's law. Accordingly, assuming that the distribution volume of EtOH contained in a 0.5 mL sample is about 2500 mL airspace at room temperature (20 °C), and that the headspace in a gas-tight sealed tube is about 1 mL when correctly filled, then the bias produced by venting once the headspace is less than -3% even at 40 °C (80,81,83,84). Therefore, a leak from the sealed cap occurring during the storage requires an extremely long period of time (*e.g.*, 1 year) to reach a bias of - 1% (85). As the loss of EtOH due to the partition in the headspace is negligible, there is no bias between 6 mL and 10 mL volume NaF/Ox pre-evacuated sterile tubes (*i.e.*, different sample-to-headspace ratio) when these are correctly filled with the whole blood (86). However, when the NaF/Ox tube is partially filled so that the NaF reaches a final concentration of 2% or 5% w/v, the bias produced at room temperature by the headspace vent is - 3.0% and - 9.0%, respectively (87). This bias is due to the "salting-out" effect of the concentrated NaF in the matrix that increases the EtOH evaporation into the headspace (81).

Remarkably, for those longer chain alcohols (*e.g.*, propyl and butyl) used as the analytical internal standard in HS-GC, the salting-out is stronger because of their naturally lower water solubility (88,89). As a consequence, in case of partial filling the EtOH/internal standard peak area ratio in the sample obtained from 1% w/v NaF/Ox tube is fictitiously reduced giving rise to bias up to - 3% if no salt-saturated sample preparation is adopted (87,90-92).

### Sample storage

In healthy subjects, it can be found no ADH in a significant concentration either within erythrocytes or free in serum (93). However, the rate of acetaldehyde formation in ethylene diamine-tetraacetic acid-anticoagulated (EDTA) whole blood containing EtOH steeply increases regardless of the addition of inhibitors (*e.g.*, citric acid, iodoacetic acid, fluoride, aminoatriazole, azide and pyrazole) of oxidases that may attack very shortchained aliphatic compounds (*e.g.*, glyceraldehyde-3-phosphate dehydrogenase, catalase, ADH) (94). Therefore, a non-enzymatic oxidation (NEO) of EtOH exists and its kinetic depends on storage temperature, matrix oxygenation and haemoglobin concentration (see Appendix B) (94-97).

Based on studies whose conditions are compatible with sampling blood from a patient using a preevacuated sterile tube (*i.e.*, blood suctioned within by the vacuum), a whole blood sample can be delivered at room temperature and stored refrigerated (*i.e.*, - 20 °C) for up to 14 days with no forensically and diagnostically significant bias (see Appendix C) (55,57,58,69,72,98). However, in case of harsh handling of the specimen before storage or analysis (*e.g.*, exposure to elevated room temperature, high thermic excursion or external transportation without refrigeration), the bias increases up to - 0.20 g/L yet within the third day from collection (56). The prolonged heating surely enhances the kinetics of the NEO, but it is also likely to cause a loosening of the gas-tight cap through which allows the venting of the tube. In fact, even a short (< 20 min) but extreme heating (burning) gives negligibly biased of the BAC if the collection tube remains intact (99).

The negative bias tends to appear by 15 days up to 3 months depending on how much the temperature, oxygenation and time are actually affecting the kinetics of NEO (see Appendix C) (95). For instance, after 1 month at - 20 °C the bias averages -6% when the headspace is 20% (i.e., 1/5 of the tube volume) (100). However, venting the tube at least once before or during storage almost doubles the bias regardless of both the temperature (by - 10 °C to + 25 °C) and the duration (days to years) of storage (55,58,60,101,102). In general, a mild storage condition (*i.e.*, + 4 °C) gives a bias that is on average within - 0.04 g/L for one-year and - 0.20 g/L for three years (103,104). As the size of the storage bias correlates well with the storage time, the correlation between storage bias and the original BAC depends on the length of the time interval over which it is computed (e.g., r = 0.80 up to 6 months, r = 0.23 up to 13 months vs. and r < 0.17 more than 5 years) (55,58,60,100).

Despite the lack of haemoglobin, the stored plasma shows - 7.8% bias after 2 months and up to -25.2% after 5 months at - 20 °C (105). This bias is reduced to - 2% after 6 months if the sample is deproteinized by acid precipitation (106). Maybe, the loss of EtOH from plasma depends on the effect of residual iron within the matrix (also freed by the acid precipitation), as well as on the integrity of the gas-tight seal at low temperature.

Finally, an effectively preserving activity of BAC is observed in carbon monoxide-saturated whole blood where no significant bias occurs due to the blockage of oxyhaemoglobin formation (107). Comparable result can be achieved by deoxygenating the whole blood with bubbling nitrogen (94). This mechanism can be relevant for preserving the BAC in the samples used for quality control and proficiency testing, and it suggests that the smoking condition might act as a preanalytical factor interacting with other factors in the bias associated with the conditions of storage (55).

#### Matrix integrity

The spectrophotometric enzymatic measurement of BAC in serum/plasma is based on the UV absorption peak of NADH + H<sup>+</sup> at 340 nm, that is produced by the ADH according to a 1:1 stoichiometry during the oxidation of EtOH (108). This absorption peak is close to that produced by free oxygenated haemoglobin. However, the modern enzymatic assays use a 20-fold dilution to measure the BAC, so that no interference (*i.e.*, positive bias) is expected even with gross haemolysis (0.8 g/dL free haemoglobin) (109). It is likely that the haemolysis bias up to - 10% is caused by the oxidation of EtOH occurring before the enzymatic analysis, for instance as the consequence of the release of catalase from the erythrocytes (109-111). As this also accelerates the depletion of the cellular antioxidants that antagonize the NEO (see Appendix B), storing a sample with haemolysis causes a bias up to - 0.03 g/L within one week (112,113).

Ethanol can bind to hydrophobic sites of the albumin via the methyl group (114). Although this interaction can displace drugs like diazepam, warfarin and corticosteroids, the affinity constant toward albumin is actually large ( $K_d = 53.1 \pm 3.1 \text{ mM}$ or  $\approx$ 2.4 g/L) (115-117). Since the formation of clots causes a minimal loss of protein and there is no sizable bound fraction of EtOH, the serum/plasma ratio of BAC is almost unity with very low variability (CV 0.01% to 0.03%) (72,118). However, since both the size and the water content (by 40% to 80% w/v) of clots is highly variable, the formation of clots within a whole blood specimen alters the distribution of water and makes unpredictable the bias caused by the inhomogeneity of the BAC (119,120). Thus, a specimen of whole blood with

clots should be homogenized before the determination of BAC but the negative bias arising from the grinding process averages - 0.01 g/L (121).

## Other relevant factors related to the control of BAC

#### Biosynthesis

In liver and kidney of humans, the mitochondrial oxidative decarboxylation of pyruvate produces acetaldehyde (122). Under hypoxic conditions, acetaldehyde is metabolized to EtOH by the cytosolic ADH in a redox reaction that converts the cofactor nicotinamide adenine dinucleotide from reduced (NADH + H<sup>+</sup>) to oxidized (NAD<sup>+</sup>) form (122). The frequency distribution of the BAC produced by this endogenous synthesis in ostensibly healthy and sober individuals is strongly left-skewed, with the 95<sup>th</sup> and 99.4<sup>th</sup> percentiles (age-, gender- and race-unrelated) corresponding to 0.012 g/L and 0.020 g/L respectively (123,124). To date, no individual value above 0.040 g/L has been reported (125).

Just in case of a rare clinical condition known as "auto-brewery" or "gut fermentation" syndrome (no more than 20 cases correctly identified in literature until 2021), usually associated with severe fungal infections, gastric resection or Crohn disease, the BAC can be as high as 0.5 g/L and up to 2.0 g/L under carbohydrates loading (126,127).

#### Biochemistry

Because of the negligible binding to proteins and membranes, the distribution of EtOH follows the water content of blood components and thus it is partitioned between the serum/plasma fraction and the cytosol of the erythrocytes (with minor contributions from platelets and leukocytes) (128). Since the average water content of erythrocytes is  $\approx$ 75% w/v (weight/volume),  $\approx$ 85% w/v of whole blood and  $\approx$ 95% w/v of serum/plasma, then the serum/whole blood as well as plasma/whole blood concentration ratio reflects the average water content of the blood components weighted by their relative abundance and is equal to 1.10-1.14 (26,72,118,128,129). This ratio varies more betweenthan within-subjects, and the CV < 3% is less than the biological variability of the haematocrit ( $CV \approx 5.5\%$ ) (26,130). Accordingly, the BAC measured in serum/plasma is unaffected by the haematocrit unless there is an extreme shift as in the polycy-thaemia or severe anaemia as during haemorrhage (73,118,128,129,131).

As water content rules the BAC, sizable deviations from the ratio of 1.10-1.14 can arise for a change in proteins and lipids content of serum/plasma (129,132). In this regard, only extreme alimentary, stress-related or genetic hyperlipidaemias can eventually reduce water content of serum by no more than 5%. On the contrary, much larger effect (in either directions) is expected for hyper- or hypoproteinaemia *via* the strong regulation of the oncotic pressure, that in turn affects the distribution of EtOH to the tissues (especially the skeletal muscle).

## Discussion

For the clinical laboratory, the BAC represents a significant organizational and cultural challenge due to its dual clinical and forensic significance. Even when not directly involved in the analysis of DUI cases - often outsourced to external laboratories for the need for specific instrumentation and expertise - the clinical laboratory may be involved in managing cases that, due to the circumstances in which they originated, can assume (often unpredictably) medico-legal relevance. It is here that the need for defensibility arises, that is, the ability to justify the forensic validity of the data, not as an added value but as an integral part of the diagnostic process. To this concern, it must be not disregarded the remarkable divide of magnitude between the allowable total error (TEa) set for the diagnostic BAC (between 20% and 9% of Clinical laboratory improvement amendments (CLIA) and Guidelines of the German Federal Medical Council (Rili-BAEK), respectively) and the combined measurement uncertainty recommended for the forensic BAC (within 4%) (133-138).

The preanalytical factors of the BAC (Figure 1) can generally be distinguished into two types: those related to the behaviour of EtOH within the patient's body, which instantaneously determine the BAC at the time of phlebotomy, and those related to the behaviour of EtOH within the blood specimen, which determine the resulting BAC after phlebotomy. Remarkably, the factors of the first type (i.e., time and site of sampling) have a limitedly controllable nature via the operative procedures (as in the case of a polytraumatized patient involved in a car accident whose exact time of last consumption is unknown). Notwithstanding their consideration falls within the logic of defensibility of the data, that for instance can be achieved through the postanalytical phase by building up an uncertainty budget derived from the knowledge of the associated bias and imprecision. The same applies to EtOH biosynthesis and biochemistry, the postanalytical control of which is perhaps the most significant expression of the dual clinical and forensic significance of the BAC, where a possible reference interval of endogenous EtOH forms the basis for a more realistic discussion on how to deal analytically with a zero-tolerance policy for EtOH consumption.

The factors of the second group (namely from sampling technique to matrix integrity), on the other hand, are operationally controllable. Therefore, they are the subject of more intense debate when the issue of coexistence between clinical and forensic preanalytics of the BAC is raised, because they determine the degree to which the operations need to be complicated to ensure the adequacy of the preanalytical process. In this regard, it should be noted that the preanalytics of the BAC is perhaps the most studied among laboratory tests and spans nearly a century of scientific publications (see the Appendix A). On the one hand, this is a favourable aspect because it indicates extensive characterization of the subject matter. On the other hand, it must be recognized that precisely because of this extensive characterization, the produced evidence has stratified through the evolution of knowledge and means that have accompanied the recent history of clinical chemistry. In other words, retrospective analysis often reveals discrepancies or incongruences among the evidence, making it difficult to grasp the correct indications provided by direct investigation or observation. In particular, this refers to the age-old

question of whether it is necessary to collect and preserve the blood sample for forensic BAC separately from diagnostic BAC, that is, whether the preanalytical processes must necessarily be distinct.

In the investigation of preanalytics, a crucial aspect is represented by the relationship established between the experimental design and the factors effectively controlled by the experimenter (139). When considering the vast amount of literature on the preanalytics of the BAC, it is possible to explain and understand this only by taking into account the impact produced by the introduction of new blood collection devices, sterile and pre-evacuated, which have been increasingly used since the 1970s. These devices are the key that translates the control of preanalytical factors, as characterized by Smalldon, Brown and colleagues in their seminal works (even though they did not use such devices), into practice (95,96). If their experimental role is not explicitly taken into consideration (which is easy because they become part of routine activity for sample collection in studies), their presence or absence in a preanalytical study acts as a confounder (see Appendix C). Supporting this consideration, it is worth noting that no preanalytics study of the BAC has ever compared the use of these new devices with the previous technique based on blood collection with a syringe and dispensing into various pre-added but open tubes.

Given the above, there is no evidence to support the need for intensifying and differentiating the use of devices and procedures for AAI or DUI cases to mitigate bias for the latter if pre-evacuated and sterile collection tubes are used according to the correct phlebotomy procedures (i.e., complete filling and mixing, vacuum sealing, and no haemolysis) (see Table 1). Indeed, under such conditions, the bias arising from intra- and extra-mural transport conditions, processing (e.g., at room temperature for < 3 hours), short-term storage (e.g., at + 4 °C for < 3 months), and long-term storage (e.g., at - 20 °C for < 3 years) adopted for diagnostics is compatible with the requirements for forensic BAC. This overall agrees with the conclusions reached by other authors through different paths, that there is no requirement for a specific preser-

Preanalytical phase of blood alcohol concentration

Ialongo C.

TABLE 1. The "seven pillars" of a diagnostically reliable and forensically defensible unified preanalytics of blood alcohol concentration

#### The "seven pillars" of a diagnostically reliable and forensically defensible unified preanalytics of BAC

- 1. use non-alcoholic antiseptics for skin cleansing
- 2. change the site of phlebotomy instead of probing/manipulating
- 3. choose either heparin or EDTA anticoagulant
- 4. use only pre-evacuated sterile collection tubes of the smallest capacity (i.e., 2.5 mL)
- 5. avoid air suction, tube venting and haemolysis before storage
- 6. freeze only for monthly or yearly storage
- 7. beware of vented or opened tubes especially for reanalysis

BAC - blood alcohol concentration. EDTA - ethylene diamine-tetra-acetic acid.

vative for forensic BAC, such as NaF (which, moreover, introduces additional complications like salting-out or haemolysis) (140,141).

The only necessary special care relates to the use of non-alcoholic disinfectants for forensic BAC, which also benefits the diagnostics of AAI. In this regard, it should be remarked that guidelines provided by official laboratory medicine bodies such as the European federation of clinical chemistry and laboratory medicine (EFLM) do not ban alcoholic antiseptics for BAC request, but advice to take adequate time for drying off the alcohol before the venipuncture (142). Therefore, the present recommendation should be regarded as a reinforcement of that guidance, in a way that makes the preparation for blood collection for BAC analysis 'error-proof' against the pressures that a challenging operational context (e.g., the Emergency Room) may exert on the correct timing of the procedure.

Since these disinfectants offer the same safety and usage procedures as alcoholic ones and are widely available in the market without significant budget impact, their substitutionary use (if not already in place) can be safely adopted for all blood sampling procedures (24). Therefore, the only necessary differentiation remains the adoption of a chain of custody for sample traceability whenever appropriate for the explicit medico-legal end of the requested BAC.

Therefore, it can be concluded that the execution of correct diagnostic preanalytical procedures guarantees both the diagnostic safety and forensic defensibility of clinical data, and no evidence contrary to a common preanalytical process can be found in the literature. It must be remarked that any deviation from these conditions magnifies the bias guaranteed by the use of pre-evacuated sterile devices (*e.g.*, opening of the tube for matrix sampling), compromising the integrity of the BAC, both forensic and diagnostic, regardless of the procedures adopted before and after its reanalysis.

#### Acknowledgments

The author wishes to thank Dr. Med. U. Wantoch von Rekowski for helping with the German papers, and Dr. Med M. Sapio for suggestions. This work is dedicated to the memory of A.M. Lostia, PhD, clever researcher and friend.

#### **Potential conflict of interest**

None declared.

#### References

- 1. World Health Organization (WHO). Global status report on alcohol and health 2018. Geneve (CH): World Health Organization; 2018. Available from https://www.who.int/ publications/i/item/9789241565639. Accessed May 5th 2023.
- 2. Griswold MG, Fullman N, Hawley C, Arian N, Zimsen SRM, Tymeson HD, et al. Alcohol use and burden for 195 countries and territories, 1990-2016: a systematic analysis for the Global Burden of Disease Study 2016. Lancet. 2018;392:1015-35. https://doi.org/10.1016/S0140-6736(18)31310-2

- 3. Centers for Disease Control and Prevention (CDC). Alcohol poisoning deaths (a deadly consequence of binge drinking) 2015. Available from https://www.cdc.gov/vitalsigns/alco-hol-poisoning-deaths/index.html. Accessed May 5th 2023.
- 4. International Traffic Safety and Analysis Group (IRTAD). Alcohol-related road casualties in official crash statistics. Paris (FR): International Transport Forum; 2017. Available from https://www.itf-oecd.org/sites/default/files/docs/ alcohol-related-road-casualties-official-crash-statistics.pdf. Accessed May 5th 2023.
- 5. European Commission (EU). Road safety thematic report Alcohol, drugs and medicine. European Road Safety Observatory. 2021. Brussels (BE). Available from: https://www.researchgate.net/publication/359921717\_European\_Road\_ Safety\_Observatory\_Road\_Safety\_Thematic\_Report\_-Alcohol\_drugs\_and\_medicine\_Road\_safety\_thematic\_report. Accessed May 5th 2023.
- 6. Rasaiah B. Blood alcohol testing. CMAJ. 1986;135:104-5.
- 7. Frajola WJ. Blood alcohol testing in the clinical laboratory: problems and suggested remedies. Clin Chem. 1993;39:377-9. https://doi.org/10.1093/clinchem/39.3.377
- 8. Faynor SM. Alcohol testing in the clinical laboratory: alternative remedies. Clin Chem. 1993;39:2539-40. https://doi. org/10.1093/clinchem/39.12.2539
- 9. Rainey PM. Alcohol testing in the clinical laboratory: alternative remedies. Clin Chem. 1993;39:2538-9; author reply 40. https://doi.org/10.1093/clinchem/39.12.2538
- 10. Urry FM, Wong Y. Current issues in alcohol testing. Lab Med. 1995;26:194-7. https://doi.org/10.1093/labmed/26.3.194
- 11. Wright JW. Alcohol and the laboratory in the United Kingdom. Ann Clin Biochem. 1991;28:212-7. https://doi. org/10.1177/000456329102800302
- 12. Gostomzyk JG, Reineke H, Neumann GK, Leithoff H. Besthet eine veno-venose rechts-links-differenz in der blutalkoholkinzentration [Is there a difference between the right and left arms concerning the alcohol concentration in the venous blood?]. Int J Legal Med. 1974;75:37-42. (in German) https://doi.org/10.1007/BF02114960
- 13. Jones AW. Pharmacokinetics of ethanol issues of forensic importance. Forensic Sci Rev. 2011;23:91-136.
- 14. Dubowski KM, Fleld PM, Frajola WJ, Raisys VA, Reeder RH, SHoemaker MJ, et al. Blood alcohol testing in the clinical laboratory; approved guideline CLSI C60 (formerly NCCLS T/ DM6-A) Wayne, PA: Clinical and Laboratory Standard Institute (CLSI); 1997. p. 44.
- 15. Springer E. Blutalkoholkurven nach Gabe von wässrigen Äthanollösungen verschiedener Konzentrationen [Blood alcohol curves after administration of aqueous ethanol solutions of various concentrations]. Blutalkohol. 1972;9:198-206. (in German)
- Ponsold A. Die Plateaubildung (Gréhant) in der Blutalkoholkurve [The plateau formation (Gréhant) in the blood alcohol curve]. In: Havard JDJ, editor. Alcohol and road traffic, proceedings of the 3rd International Conference on alcohol and road traffic. London (UK): British Medical Association; 1963. p. 197-202.
- 17. Jones AW, Jonsson KA, Jorfeldt L. Differences between capillary and venous blood-alcohol concentrations as a func-

tion of time after drinking, with emphasis on sampling variations in left vs right arm. Clin Chem. 1989;35(3):400-4. https://doi.org/10.1093/clinchem/35.3.400

- McCallum NEW, Scroggie JG. Some aspects of alcohol in body fluids. III. Study of alcohol in blood in different parts of the body. Med J Aust. 1960;2:1031-2. https://doi. org/10.5694/j.1326-5377.1960.tb24039.x
- 19. Sedman AJ, Wilkinson PK, Wagner JG. Concentrations of ethanol in two segments of the vascular system. J Forensic Sci. 1976;21:315-22. https://doi.org/10.1520/JFS10499J
- 20. Jones AW, Lindberg L, Olsson SG. Magnitude and time-course of arterio-venous differences in blood-alcohol concentration in healthy men. Clin Pharmacokinet. 2004;43:1157-66. https://doi.org/10.2165/00003088-200443150-00006
- 21. Harger RN. Blood source and alcohol level: errors from using venous blood during active absorption. In: Havard JDJ, editor. Alcohol and road traffic, proceedings of the 3rd International Conference on alcohol and road traffic. London, UK: British Medical Association; 1963. p. 212-9.
- 22. Jones AW, Jorfeldt L, Hjertberg H, Jonsson KA. Physiological variations in blood ethanol measurements during the postabsorptive state. J Forensic Sci Soc. 1990;30:273-83. https:// doi.org/10.1016/S0015-7368(90)73355-5
- 23. Taylor L, Remeškevičius V, Saskoy L, Brodie T, Mahmud J, Moir H. Determination of ethanol in micro-volumes of blood by headspace gas chromatography: Statistical comparison between capillary and venous sampling sites. Med Sci Law. 2021;61:86-96. https://doi.org/10.1177/0025802420928632
- 24. Ialongo C, Bernardini S. Phlebotomy, a bridge between laboratory and patient. Biochem Med (Zagreb). 2016;26:17-33. https://doi.org/10.11613/BM.2016.002
- 25. Naeve W, Koops E, Audrlický I, Brinkmann B. Veno-venöse Differenzen der Blutalkoholkonzentration nach "Sturztrunk" [Veno-venous differences in blood alcohol concentration after "heavy drink"]. Blutalkohol. 1971;8. (in German)
- 26. Jones AW, Hahn RG, Stalberg HP. Distribution of ethanol and water between plasma and whole blood; inter- and intra-individual variations after administration of ethanol by intravenous infusion. Scand J Clin Lab Invest. 1990;50:775-80. https://doi.org/10.1080/00365519009091072
- 27. Norberg A, Gabrielsson J, Jones AW, Hahn RG. Within- and between-subject variations in pharmacokinetic parameters of ethanol by analysis of breath, venous blood and urine. Br J Clin Pharmacol. 2000;49:399-408. https://doi.org/10.1046/ j.1365-2125.2000.00194.x
- 28. Norberg A, Jones AW, Hahn RG. Pharmacokinetics of ethanol in arterial and venous blood and in end-expired breath during vasoconstriction and vasodilation. Am J Ther. 1995;2:954-61. https://doi.org/10.1097/00045391-199512000-00009
- 29. Zink P, Blauth M. Zur Frage der Beenflussung der Alkoholkonzentration im Cubitalvenenblut durch die Blutentnahmetechnik [The influence of the tecnique of taking blood from the cubital vein on the blood ethanol concentration]. Blutalkohol. 1982;19:75-87. (in German)
- 30. Pendlington RU, Whittle E, Robinson JA, Howes D. Fate of ethanol topically applied to skin. Food Chem Toxi-

col. 2001;39:169-74. https://doi.org/10.1016/S0278-6915(00)00120-4

- 31. Dubowski KM, Essary NA. Contamination of blood specimens for alcohol analysis during collection. Abstr Rev Alcohol Driving. 1983;4:3-8.
- 32. Winek CL, Eastly T. Factors affecting contamination of blood samples for ethanol determinations. Leg Med Annu. 1977;1976:147-62.
- 33. Müller FO, Hundt HK. Letter: Ethyl alcohol: contamination of blood specimens. S Afr Med J. 1976;50:91.
- 34. Lippi G, Simundic AM, Musile G, Danese E, Salvagno G, Tagliaro F. The alcohol used for cleansing the venipuncture site does not jeopardize blood and plasma alcohol measurement with head-space gas chromatography and an enzymatic assay. Biochem Med (Zagreb). 2017;27:398-403. https://doi.org/10.11613/BM.2017.041
- 35. Heise HA. How extraneous alcohol affects the blood test for alcohol; pitfalls to be avoided when withdrawing blood for medicolegal purposes. Am J Clin Pathol. 1959;32:169-70. https://doi.org/10.1093/ajcp/32.2.169
- 36. Higuchi A, Kurihara R, Yoshimoto T, Sano T, Katsumata K, Ishii A, et al. Problems in blood alcohol testing of severely injured drivers brought to emergency departments in Japan. Leg Med (Tokyo). 2005;7:299-305. https://doi.org/10.1016/j. legalmed.2005.05.003
- 37. McIvor RA, Cosbey SH. Effect of using alcoholic and non-alcoholic skin cleansing swabs when sampling blood for alcohol estimation using gas chromatography. Br J Clin Pract. 1990;44:235-6. https://doi.org/10.1111/j.1742-1241.1990. tb10800.x
- 38. Peek GJ, Marsh A, Keating J, Ward RJ, Peters TJ. The effects of swabbing the skin on apparent blood ethanol concentration. Alcohol Alcohol. 1990;25:639-40. https://doi. org/10.1093/oxfordjournals.alcalc.a045060
- 39. Carter PG, McConnell AA. Alcohol in drinking driving swabs: does it make any difference? Med Sci Law. 1990;30:90.
- 40. Peek GJP, Keating JW, Ward RJ, Peters TJ. Alcohol swabs and venepuncture. The Lancet. 1989;17:1388. https://doi. org/10.1016/S0140-6736(89)92830-4
- 41. Tucker A, Trethewy C. Lack of effect on blood alcohol level of swabbing venepuncture sites with 70% isopropyl alcohol. Emerg Med Australas. 2010;22:9-12. https://doi. org/10.1111/j.1742-6723.2009.01248.x
- 42. Miller MA, Rosin A, Levsky ME, Gregory TJ, Crystal CS. Isopropyl alcohol pad use for blood ethanol sampling does not cause false-positive results. J Emerg Med. 2007;33:9-10. https://doi.org/10.1016/j.jemermed.2006.10.005
- 43. Levsky ME, Miller MA. Isopropyl alcohol skin prep pads: the extreme case. J Emerg Med. 2007;33:289. https://doi. org/10.1016/j.jemermed.2007.02.063
- 44. Taberner PV. A source of error in blood alcohol analysis. Alcohol and Alcoholism. 1989;24:489-90.
- 45. Grice EA, Segre JA. The skin microbiome. Nat Rev Microbiol. 2011;9:244-53. https://doi.org/10.1038/nrmicro2537
- 46. Chen YE, Tsao H. The skin microbiome: current perspectives and future challenges. J Am Acad Dermatol. 2013;69:143-55. https://doi.org/10.1016/j.jaad.2013.01.016

- 47. Findley K, Oh J, Yang J, Conlan S, Deming C, Meyer JA, et al. Topographic diversity of fungal and bacterial communities in human skin. Nature. 2013;498:367-70. https://doi. org/10.1038/nature12171
- 48. Corry JE. A review. Possible sources of ethanol ante- and post-mortem: its relationship to the biochemistry and microbiology of decomposition. J Appl Bacteriol. 1978;44:1-56. https://doi.org/10.1111/j.1365-2672.1978.tb00776.x
- 49. Mayser P, Fuhrer D, Schmidt R, Grunder K. Hydrolysis of fatty acid esters by Malassezia furfur: different utilization depending on alcohol moiety. Acta Derm Venereol. 1995;75:105-9. https://doi.org/10.2340/0001555575105109
- 50. Dotzauer G, Redetzki H, Johannsmeier K, Bücher T. Erprobung einer spezifischen Fermentmethode zur Mikrobestimmung von Äthylalkohol [Investigastions on a specific fermentation method for the micro determination of ethyl alcohol]. Int J Legal Med. 1952;41:15-29.(in German) https:// doi.org/10.1007/BF00664479
- 51. Laurens JB, Sewell FJJ, Kock MM. Pre-analytical factors related to the stability of ethanol concentration during storage of ante-mortem blood alcohol specimens. J Forensic Leg Med. 2018;58:155-63. https://doi.org/10.1016/j. jflm.2018.06.003
- 52. Dick GL, Stone HM. Alcohol loss arising from microbial contamination of drivers' blood specimens. Forensic Sci Int. 1987;34:17-27. https://doi.org/10.1016/0379-0738(87)90080-6
- 53. Yajima D, Motani H, Kamei K, Sato Y, Hayakawa M, Iwase H. Ethanol production by Candida albicans in postmortem human blood samples: effects of blood glucose level and dilution. Forensic Sci Int. 2006;164:116-21. https://doi. org/10.1016/j.forsciint.2005.12.009
- 54. Blume P, Lakatua DJ. The effect of microbial contamination of the blood sample on the determination of ethanol levels in serum. Am J Clin Pathol. 1973;60:700-2. https://doi. org/10.1093/ajcp/60.5.700
- 55. Jones AW. Are changes in blood-ethanol concentration during storage analytically significant? Importance of method imprecision. Clin Chem Lab Med. 2007;45:1299-304. https:// doi.org/10.1515/CCLM.2007.289
- 56. Glover PL. The effect of heat on blood samples containing alcohol. 2002 Conference of the International Association for Chemical Testing; Austin (TX). 2002.
- 57. Vance CS, Carter CR, Carter RJ, Del Valle MM, Pena JR. Comparison of immediate and delayed blood alcohol concentration testing. J Anal Toxicol. 2015;39:538-44. https://doi. org/10.1093/jat/bkv061
- 58. Shan X, Tiscione NB, Alford I, Yeatman DT. A study of blood alcohol stability in forensic antemortem blood samples. Forensic Sci Int. 2011;211:47-50. https://doi.org/10.1016/j.forsciint.2011.04.012
- 59. Petković S, Savic S, Zgonjanin D, Samojlik I. Ethanol concentrations in antemortem blood samples under controlled conditions. Alcohol Alcohol. 2008;43:658-60. https://doi. org/10.1093/alcalc/agn072
- 60. Tiscione NB, Vacha RE, Alford I, Yeatman DT, Shan X. Longterm blood alcohol stability in forensic antemortem whole

blood samples. J Anal Toxicol. 2015;39:419-25. https://doi. org/10.1093/jat/bkv037

- 61. Kaye S, Dammin GJ. Stability of the blood alcohol: an agent to maintain the alcohol concentration in drawn blood. Mil Surg. 1945;96:93-5. https://doi.org/10.1093/milmed/96.1.93
- 62. Li S, Smith KD, Davis JH, Gordon PB, Breaker RR, Strobel SA. Eukaryotic resistance to fluoride toxicity mediated by a widespread family of fluoride export proteins. Proc Natl Acad Sci U S A. 2013;110:19018-23. https://doi.org/10.1073/ pnas.1310439110
- 63. Sangster J. Octanol-water partition coefficients of simple organic compounds. J Phys Chem Ref Data. 1989;18:1111-27. https://doi.org/10.1063/1.555833
- 64. Bowen RAR, Remaley AT. Interferences from blood collection tube components on clinical chemistry assays. Biochem Med (Zagreb). 2014;24:31-44. https://doi.org/10.11613/ BM.2014.006
- 65. Steuer C, Huber AR, Bernasconi L. Where clinical chemistry meets medicinal chemistry. Systematic analysis of physico-chemical properties predicts stability of common used drugs in gel separator serum tubes. Clin Chim Acta. 2016;462:23-7. https://doi.org/10.1016/j.cca.2016.08.014
- 66. Schrapp A, Mory C, Duflot T, Pereira T, Imbert L, Lamoureux F. The right blood collection tube for therapeutic drug monitoring and toxicology screening procedures: Standard tubes, gel or mechanical separator? Clin Chim Acta. 2019;488:196-201. https://doi.org/10.1016/j.cca.2018.10.043
- 67. Laessig RH, Westgard JO, Carey RN, Hassemer DJ, Schwartz TH, Feldbruegge DH. Assessment of a serum separator device for obtaining serum specimens suitable for clinical analyses. Clin Chem. 1976;22:235-9. https://doi. org/10.1093/clinchem/22.2.235
- 68. Dasgupta A, Yared MA, Wells A. Time-dependent absorption of therapeutic drugs by the gel of the Greiner Vacuette blood collection tube. Ther Drug Monit. 2000;22:427-31. https://doi.org/10.1097/00007691-200008000-00011
- 69. Karinen R, Oiestad EL, Andresen W, Wethe G, Smith-Kielland A, Christophersen A. Comparison of ethanol and other drugs of abuse concentrations in whole blood stored in venoject glass and plastic and venosafe plastic evacuated tubes. J Anal Toxicol. 2010;34:420-8. https://doi.org/10.1093/ jat/34.7.420
- 70. Jones AW, Ericsson E. Decreases in blood ethanol concentrations during storage at 4 degrees C for 12 months were the same for specimens kept in glass or plastic tubes. Pract Lab Med. 2016;4:76-81. https://doi.org/10.1016/j. plabm.2016.02.002
- 71. Bowen RA, Adcock DM. Blood collection tubes as medical devices: the potential to affect assays and proposed verification and validation processes for the clinical laboratory. Clin Biochem. 2016;49:1321-30. https://doi.org/10.1016/j. clinbiochem.2016.10.004
- 72. Penetar DM, McNeil JF, Ryan ET, Lukas SE. Comparison among plasma, serum, and whole blood ethanol concentrations: impact of storage conditions and collection tubes. J Anal Toxicol. 2008;32:505-10. https://doi.org/10.1093/ jat/32.7.505

- 73. Wilkinson PK, Wagner JG, Sedman AJ. Sensitive head-space gas chromatographic method for the determination of ethanol utilizing capillary blood samples. Anal Chem. 1975;47:1506-10. https://doi.org/10.1021/ac60359a048
- 74. Streete PJ, Flanagan RJ. Ethylbenzene and xylene from Sarstedt Monovette serum gel blood-collection tubes. Clin Chem. 1993;39:1344-5. https://doi.org/10.1093/ clinchem/39.6.1344
- 75. Dyne D, Cocker J, Streete PJ, Flanagan RJ. Toluene, 1-butanol, ethylbenzene and xylene from Sarstedt Monovette serum gel blood collection tubes. Ann Clin Biochem. 1996;33:355-6. https://doi.org/10.1177/000456329603300414
- 76. Tarnovski G, Hayashi T, Igarashi K, Ochi H, Matoba R. Misidentification of ethyl chloride in the routine GC-FID analysis for alcohol. Forensic Sci Int. 2009;188:e7-9. https://doi. org/10.1016/j.forsciint.2009.02.008
- Laferty PI. Ethyl chloride: possible misidentification as ethanol. J Forensic Sci. 1994;39:261-5. https://doi.org/10.1520/ JFS13595J
- 78. Kosecki PA, Autret A, Abbott L, Keller-Brooke K. Isobutylene contamination of blood collected in 10-ml evacuated blood collection tubes with gray conventional rubber stoppers. J Forensic Sci. 2021;66:2484-92. https://doi. org/10.1111/1556-4029.14792
- 79. Boumba V. Impact of blood collection tubes on erroneous 1-propanol detection and on forensic ethanol analysis. J Forensic Toxicol Pharmacol. 2015;4:1. https://doi. org/10.4172/2325-9841.1000134
- Harger RN, Raney BB, Bridwell EG, Kitchel MF. The partition ratioof alcohol between air and water, urine and blood; estimation and identification of alcohol in these liquids from analysis of air equilibrated with them. J Biol Chem. 1950;197-213. https://doi.org/10.1016/S0021-9258(18)56458-9
- 81. Jones AW. Determination of liquid/air partition coefficients for dilute solutions of ethanol in water, whole blood, and plasma. J Anal Toxicol. 1983;7:193-7. https://doi.org/10.1093/jat/7.4.193
- 82. Saracevic A, Simundic AM, Dukic L. The stability of ethanol in unstoppered tubes. Clin Biochem. 2014;47:92-5. https:// doi.org/10.1016/j.clinbiochem.2013.11.006
- 83. Karger J, Sachs V. Über die Ursachen des Konzentrationsabfalles in ordnungsgemäß verschlossenen und aufbewahrten alkoholhaltigen Blutproben [On the causes of concentration drop in properly sealed and stored alcoholic blood samples]. Int J Legal Med. 1958;47:614-8. (in German) https://doi.org/10.1007/BF00664873
- 84. Bucher T, Redetzki H. Eine spezifische photometrische bestimmung von athylalkohol auf fermentativem wege [A specific photometric determination of ethyl alcohol by fermentation]. Klin Wochenschr. 1951;29:615-6. (in German) https://doi.org/10.1007/BF01485653
- 85. Sachs V. Zur Frage der Verdunstung von Alkohol in ordnungsgemäß verschlossenen Blutproben [On the question of evaporation of alcohol in properly sealed blood samples]. Int J Legal Med. 1959;48:400-2. (in German) https://doi. org/10.1007/BF00666946

- 86. Olds ML, Naquin JL. Statistical comparisons of blood alcohol samples from 6-mL and 10-mL grey-top tubes. J Forensic Sci. 2021;66:687-93. https://doi.org/10.1111/1556-4029.14632
- 87. Miller BA, Day SM, Vasquez TE, Evans FM. Absence of salting out effects in forensic blood alcohol determination at various concentrations of sodium fluoride using semi-automated headspace gas chromatography. Sci Justice. 2004;44:73-6. https://doi.org/10.1016/S1355-0306(04)71691-3
- 88. Saigal S, Watts J, Campbell D. Randomized clinical trial of an oscillating air mattress in preterm infants: effect on apnea, growth, and development. J Pediatr. 1986;109:857-64. https://doi.org/10.1016/S0022-3476(86)80714-4
- 89. Watts MT, McDonald OL. The effect of sodium chloride concentration, water content, and protein on the gas chromatographic headspace analysis of ethanol in plasma. Am J Clin Pathol. 1990;93:357-62. https://doi.org/10.1093/ ajcp/93.3.357
- 90. Jones AW. Salting-out effect of sodium fluoride and its influence on the analysis of ethanol by headspace gas chromatography. J Anal Toxicol. 1994;18:292-3. https://doi. org/10.1093/jat/18.5.292
- 91. Solanky AA. Effect of different concentrations of sodium fluoride on blood alcohol determination by headspace gas chromatography using the internal standard method. J Anal Toxicol. 1994;18:63. https://doi.org/10.1093/ jat/18.1.63
- 92. Jones AW, Fransson M. Blood analysis by headspace gas chromatography: does a deficient sample volume distort ethanol concentration? Med Sci Law. 2003;43:241-7. https:// doi.org/10.1258/rsmmsl.43.3.241
- 93. Tyulina OV, Prokopieva VD, Dodd RD, Hawkins JR, Clay SW, Wilson DO, et al. In vitro effects of ethanol, acetaldehyde and fatty acid ethyl esters on human erythrocytes. Alcohol Alcohol. 2002;37:179-86. https://doi.org/10.1093/alcalc/37.2.179
- 94. Chen HM, Lin WW, Ferguson KH, Scott BK, Peterson CM. Studies of the oxidation of ethanol to acetaldehyde by oxyhemoglobin using fluorigenic high-performance liquid chromatography. Alcohol Clin Exp Res. 1994;18:1202-6. https:// doi.org/10.1111/j.1530-0277.1994.tb00105.x
- 95. Brown GA, Neylan D, Reynolds WJ, Smalldon KW. The stability of ethanol in stored blood. I. Important variables and interpretation of results. Anal Chim Acta. 1973;66:271-83. https://doi.org/10.1016/S0003-2670(01)82539-3
- 96. Smalldon KW, Brown GA. The stability of ethanol in stored blood. II. The mechanism of ethanol oxidation. Anal Chim Acta. 1973;66:285-90. https://doi.org/10.1016/S0003-2670(01)82540-X
- 97. Winek T, Winek CL, Wahba WW. The effect of storage at various temperatures on blood alcohol concentration. Forensic Sci Int. 1996;78:179-85. https://doi.org/10.1016/0379-0738(95)01884-0
- 98. Winek CL, Paul LJ. Effect of short-term storage conditions on alcohol concentrations in blood from living human subjects. Clin Chem. 1983;29:1959-60. https://doi.org/10.1093/clinchem/29.11.1959

- 99. Kosecki PA, Canonico E, Brooke P. Testing antemortem blood for ethanol concentration from a blood kit in a refrigerator fire. J Forensic Sci. 2020;65:2198-200. https://doi. org/10.1111/1556-4029.14510
- 100. Stojiljkovic G, Maletin M, Stojic D, Brkic S, Abenavoli L. Ethanol concentration changes in blood samples during medium-term refrigerated storage. Eur Rev Med Pharmacol Sci. 2016;20:4831-6.
- 101. Chang RB, Smith WA, Walkin E, Reynolds PC. The stability of ethyl alcohol in forensic blood specimens. J Anal Toxicol. 1984;8:66-7. https://doi.org/10.1093/jat/8.2.66
- 102. Ferrari LA, Triszcz JM, Giannuzzi L. Kinetics of ethanol degradation in forensic blood samples. Forensic Sci Int. 2006;161:144-50. https://doi.org/10.1016/j.forsciint.2006.02.049
- 103. Kosecki PA, Raines ME. Testing antemortem blood samples for ethanol after four to seven years of refrigerated storage. J Forensic Sci. 2022;67:1250-7. https://doi. org/10.1111/1556-4029.14984
- 104. Kosecki PA, Abbott LA, Raines ME. Large-scale reanalysis of refrigerated antemortem blood samples for ethanol content at random intervals. J Forensic Sci. 2021;66:1966-72. https://doi.org/10.1111/1556-4029.14796
- 105. Kocak FE, Isiklar OO, Kocak H, Meral A. Comparison of blood ethanol stabilities in different storage periods. Biochem Med (Zagreb). 2015;25:57-63. https://doi. org/10.11613/BM.2015.006
- 106. Meyer T, Monge PK, Sakshaug J. Storage of blood samples containing alcohol. Acta Pharmacol Toxicol (Copenh). 1979;45:282-6. https://doi.org/10.1111/j.1600-0773.1979. tb02394.x
- 107. Dubowski KM, Gadsden R, Poklis A. The stability of ethanol in human whole blood controls: an interlaboratory evaluation. J Anal Toxicol. 1997;21:486-91. https://doi. org/10.1093/jat/21.6.486
- 108. Bonnichsen RK, Theorell H. An enzymatic method for the microdetermination of ethanol. Scand J Clin Lab Invest. 1951;3:58-62. https://doi. org/10.3109/00365515109060572
- 109. Ji JZ, Meng QH. Evaluation of the interference of hemoglobin, bilirubin, and lipids on Roche Cobas 6000 assays. Clin Chim Acta. 2011;412:1550-3. https://doi. org/10.1016/j.cca.2011.04.034
- 110. Lippi G, Mercadanti M, Musa R, Aloe R, Cervellin G. The concentration of plasma ethanol measured with an enzymatic assay is decreased in hemolyzed specimens. Clin Chim Acta. 2012;413:356-7. https://doi.org/10.1016/j. cca.2011.10.002
- 111. Çat A, Uçar KT, Gümüş A. Effect of haemolysis on an enzymatic measurement of ethanol. Biochem Med (Zagreb). 2021;31:010704. https://doi.org/10.11613/ BM.2021.010704
- 112. Kristoffersen L, Stormyhr LE, Smith-Kielland A. Headspace gas chromatographic determination of ethanol: the use of factorial design to study effects of blood storage and headspace conditions on ethanol stability and acetaldehyde formation in whole blood and plasma. Foren-

sic Sci Int. 2006;161:151-7. https://doi.org/10.1016/j.forsciint.2006.03.034

- 113. Kosecki PA, Brooke P, Abbott L, Canonico E. The effect of sample hemolysis on blood ethanol analysis using head-space gas chromatography. J Forensic Sci. 2021;66:1136-42. https://doi.org/10.1111/1556-4029.14675
- 114. Avdulov NA, Chochina SV, Daragan VA, Schroeder F, Mayo KH, Wood WG. Direct binding of ethanol to bovine serum albumin: a fluorescent and 13C NMR multiplet relaxation study. Biochemistry. 1996;35:340-7. https://doi. org/10.1021/bi9513416
- 115. Sandor P, Naranjo CA, Khouw V, Sellers EM. Variations in drug free fraction during alcohol withdrawal. Br J Clin Pharmacol. 1983;15:481-6. https://doi. org/10.1111/j.1365-2125.1983.tb01533.x
- 116. Hiramatsu R, Nisula BC. Effect of alcohol on the interaction of cortisol with plasma proteins, glucocorticoid receptors and erythrocytes. J Steroid Biochem. 1989;33:65-70. https://doi.org/10.1016/0022-4731(89)90359-2
- 117. Ha C-E, Petersen CE, Park DS, Harohalli K, Bhagavan NV. Investigations of the effects of ethanol on warfarin binding to human serum albumin. J Biomed Sci. 2000;7:114-21. https://doi.org/10.1007/BF02256617
- 118. Winek CL, Carfagna M. Comparison of plasma, serum, and whole blood ethanol concentrations. J Anal Toxicol. 1987;11:267-8. https://doi.org/10.1093/jat/11.6.267
- 119. Illchmann-Christ A. Untersuchungen über die Relationen von Blutkuchen-Vollblut-(Serum-)Alkoholwerten [Studies on the bloot clot-whole blood-(serum) alcohol levels]. Int J Legal Med. 1959;49:113-29. (in German) https://doi. org/10.1007/BF00665128
- 120. Hallermann W, Sachs V, Steigleder E. Alkoholbestimmungen in geronnenem Blut [Alcohol determinations in clotted blood]. Int J Legal Med. 1960;49:431-40. (in German) https://doi.org/10.1007/BF00667759
- 121. Senkowski CM, Thompson KA. The accuracy of blood alcohol analysis using headspace gas chromatography when performed on clotted samples. J Forensic Sci. 1990;35:176-80. https://doi.org/10.1520/JFS12814J
- 122. Ostrovsky YuM. Endogenous ethanol—Its metabolic, behavioral and biomedical significance. Alcohol. 1986;3:239-47. https://doi.org/10.1016/0741-8329(86)90032-7
- 123. Al-Awadhi A, Wasfi IA, Al Reyami F, Al-Hatali Z. Autobrewing revisited: endogenous concentrations of blood ethanol in residents of the United Arab Emirates. Sci Justice. 2004;44:149-52. https://doi.org/10.1016/S1355-0306(04)71707-4
- 124. Ragab AR, Al-Mazroua MK, Afify MM, Saeed IA, Katbai C. Endogenous Ethanol Production Levels in Saudi Arabia Residents. J Alcohol Drug Depend. 2015;3:211. https:// doi.org/10.4172/2329-6488.1000211
- 125. Wittig H, Böttcher S, Römhild W, Bartels H, Krause D, Jachau K, editors. Endogener Alkohol als mögliche Schutzbehauptung nach Einführung eines absoluten Alkoholverbotes für FahranfängerInnen [Endogenous alcohol as a possible protective claim after the introduction of an absolute ban on alcohol for novice drivers]. Kongressberich 2007 der Deutschen Gesellschaft fur Verkehrsme-

dizin (DGVM eV:). 2008.Heidelberg (DE). Available from https://bast.opus.hbz-nrw.de/frontdoor/index/index/ year/2016/docld/1660. Accessed May 5th 2023.

- 126. Bayoumy AB, Mulder CJJ, Mol JJ, Tushuizen ME. Gut fermentation syndrome: a systematic review of case reports. United Eur Gastroenterol J. 2021;9(3):332-42. https://doi. org/10.1002/ueg2.12062
- 127. Dinis-Oliveira RJ. The auto-brewery syndrome: a perfect metabolic "storm" with clinical and forensic implications. J Clin Med. 2021;10:4637. https://doi.org/10.3390/ jcm10204637
- 128. Rainey PM. Relation between serum and whole-blood ethanol concentrations. Clin Chem. 1993;39:2288-92. https://doi.org/10.1093/clinchem/39.11.2288
- 129. Charlebois RC, Corbett MR, Wigmore JG. Comparison of ethanol concentrations in blood, serum, and blood cells for forensic application. J Anal Toxicol. 1996;20:171-8. https://doi.org/10.1093/jat/20.3.171
- 130. Coşkun A, Carobene A, Kilercik M, Serteser M, Sandberg S, Aarsand AK, et al. Within-subject and between-subject biological variation estimates of 21 hematological parameters in 30 healthy subjects. Clin Chem Lab Med. 2018;56:1309-18. https://doi.org/10.1515/cclm-2017-1155
- 131. Shajani NK, Godolphin W, Image BA. Blood alcohol analysis: comparison of whole blood analysis by gas chromatography with serum analysis by enzymatic method. J Can Soc Forensic Sci. 1989;22:317-20. https://doi.or g/10.1080/00085030.1989.10757441
- 132. Brettel H-F. Hyperlipidämie und Blutalkoholbestimmung [Hyperlipidemia and blood alcohol determination]. Int J Legal Med. 1974;74:181-5. (in German) https://doi. org/10.1007/BF00200587
- 133. Bundesarztekammer (German Medical Association) leV. Richtlinie der Bundesarztekammer zur Qualitatssicherung laboraoriumsmedizinischer Untersuchungen (Guidelines of the German Medical Association onquality assurance in medical laboratory testing - Rili-BAEK) GMS Z Forder Qualitatssich Med Lab. 2015;6:1-43.
- 134. Heitler C, Scaife DB, Thompson BW. The oxidation of ethanol by hydrogen peroxide. Part I. Catalysis by ferric ion. J Chem Soc A. 1967:1409-13. https://doi.org/10.1039/ j19670001409
- 135. (SWGTOX) SWGFFT. Standard practices for method validation in forensic toxicology. J Anal Toxicol. 2013;37:452-74. https://doi.org/10.1093/jat/bkt054
- 136. Sklerov JH, Couper FJ. Calculation and verification of blood ethanol measurement uncertainty for headspace gas chromatography. J Anal Toxicol. 2011;35:402-10. https:// doi.org/10.1093/anatox/35.7.402
- 137. Gullberg RG. Estimating the measurement uncertainty in forensic blood alcohol analysis. J Anal Toxicol. 2012;36:153-61. https://doi.org/10.1093/jat/bks012
- 138. Hwang R-J, Beltran J, Rogers C, Barlow J, Razatos G. Measurement of uncertainty for blood alcohol concentration by headspace gas chromatography. J Can Soc Forensic Sci. 2017;50:114-24. https://doi.org/10.1080/0008503 0.2017.1312069

- 139. Ialongo C, Bernardini S. Preanalytical investigations of phlebotomy: methodological aspects, pitfalls and recommendations. Biochem Med (Zagreb). 2017;27:177-91. https://doi.org/10.11613/BM.2017.020
- 140. Kaye S. The collection and handling of the blood alcohol specimen. Am J Clin Pathol. 1980;74:743-6. https://doi. org/10.1093/ajcp/74.5.743
- 141. Rodda LN, Pearring S, Harper CE, Tiscione NB, Jones AW. Inferences and legal considerations following a blood collection tube recall. J Anal Toxicol. 2021;45:211-4. https://doi.org/10.1093/jat/bkaa056
- 142. Simundic AM, Bölenius K, Cadamuro J, Church S, Cornes MP, van Dongen-Lases EC, et al.; Working Group for Preanalytical Phase (WG-PRE), of the European Federation of Clinical Chemistry and Laboratory Medicine (EFLM) and Latin American Working Group for Preanalytical Phase (WG-PRE-LATAM) of the Latin America Confederation of Clinical Biochemistry (COLABIOCLI). Joint EFLM-COLABIOCLI Recommendation for venous blood sampling. Clin Chem Lab Med. 2018;56:2015-38. https://doi. org/10.1515/cclm-2018-0602
- 143. Marron T, Hilbe JJ, editors. A study of preserved blood specimens taken for alcohol determination. Proc Iowa Acad Sci. 1940;47:225-9.
- 144. Glendening BL, Waugh T. The stability of ordinary blood alcohol samples held various periods of time under different conditions. J Forensic Sci. 1965;10:192-200.
- 145. Jones A. Fifty years of Blutalkohol an appreciation from Sweden. Blutalkohol. 2011;48:309-17.

- 146. Winterbourn CC. Toxicity of iron and hydrogen peroxide: the Fenton reaction. Toxicol Lett. 1995;82-83:969-74. https://doi.org/10.1016/0378-4274(95)03532-X
- 147. Balagopalakrishna C, Manoharan PT, Abugo OO, Rifkind JM. Production of superoxide from hemoglobinbound oxygen under hypoxic conditions. Biochemistry. 1996;35:6393-8. https://doi.org/10.1021/bi952875+
- 148. Rifkind JM, Ramasamy S, Manoharan PT, Nagababu E, Mohanty JG. Redox reactions of hemoglobin. Antioxid Redox Signal. 2004;6:657-66. https://doi. org/10.1089/152308604773934422
- 149. Bonnichsen R, Lundgren G. Comparison of the ADH and the Widmark procedures in forensic chemistry for determinating alcohol. Acta Pharmacol Toxicol (Copenh). 1957;13:256-66. https://doi. org/10.1111/j.1600-0773.1957.tb00262.x
- 150. Krauland W, Vidic E, Freudenberg K, Schmidt B, Lenk VS. Über den Beweiswert einer zweiten Blutalkoholbestimmung an länger gelagerten Blutproben [On the probative value of a second blood alcohol determination on blood samples that have been stored for a long time]. Dtsch Z Gesam Gerichtl Med. 1959;50:34-53. (in German) https://doi.org/10.1007/BF00573690
- 151. Krauland W, Vidic E, Freudenberg K. Über den Beweiswert einer zweiten Blutalkoholbestimmung an länger gelagerten Blutproben II. Mitteilung [On the causes of the drop in concentration in properly sealed and stored blood samples containing alcohol II. Communication]. Dtsch Z Gesam Gerichtl Med. 1961;52:76-89. (in German) https://doi. org/10.1007/BF00597182

## Appendix A – a historical note on the studies on the preanalytical phase of blood alcohol concentration

In the body of studies that make up the scientific literature on blood alcohol concentration (BAC) preanalytics, it is difficult to make a clear separation between the contributions of biochemical-clinical, forensic and pharmacological research. However, there is no doubt that forensic medicine has given a fundamental impetus to this field, certainly long before preanalytics was recognized as a subject of study in its own right within laboratory medicine. The systematic study of the preparation and storage of the sample, in fact, represents the way (documented since 1940) to give a scientifically authoritative answer to the disputes about the BAC measurement results often raised in cases of drunk driving (35,61,143,144).

As evident, this review is distinguished by the citation of a large number of articles originally written in German, many of which represent a unique source of evidence on certain aspects of the preanalytics of the BAC (see for instance references (119,120)). The journals that have published many of these articles, such as "Blutalkohol" (ISSN: 0006-5250, https://www.bads.de/wissen/fachzeitschrift-blutalkohol/) and "Deutsche Zeitschrift für die gesamte gerichtliche Medizin" (in later "Zeitschrift für Rechtsmedizin" and currently "International Journal of Legal Medicine"), played a leading scientific role in studies on the determination of BAC during the second half of the 21st century (145). It is therefore not surprising that the British Medical Association's "Proceedings of the International Conference on Alcohol and Road Traffic" included scientific contributions written in German without the need for English translation.

### Appendix B – the mechanism of the non-enzymatic oxidation (NEO) of ethanol

The biochemical details of the non-enzymatic oxidation (NEO) are still unclear albeit there are several clues to the putative mechanism. First, the energy barrier to the formation of acetaldehyde (i.e., activation energy) is only relatively higher in ante-mortem (≈25 kcal/mol) than in post-mortem (≈12 Kcal/mol) samples (95,102). This suggests that no enzyme is involved, even because the activation energy decreases almost tenfold when the headspace changes from practically zero to about 2/3 of the tube volume (102). Second, the kinetics of reaction respect to the ethanol is first or pseudo-first order (i.e., concentration dependent) in post-mortem whilst it is zero-order (i.e., concentration independent) in ante-mortem blood samples (55,95,96,102). Thus, the oxidation rate is constrained by the counteracting anti-oxidant defenses that are operating in the viable erythrocytes. Third, those agents that directly reduce the haeme (azide and nitrite) or block its oxidation (cyanide) or deoxygenate the matrix (nitrogen bubbling), abolish almost completely the oxidation of ethanol in whole blood even at room temperature (94,96). Fourth, the haematin alone is able to oxidize the ethanol although to a lower rate than the intact haemoglobin (94). Fifth, the oxidation rate is dependent on temperature (112).

Collectively, the evidences suggest that the NEO follows a Fenton catalysis mediated by haeme and involving hydrogen peroxide  $(H_2O_2)$  (134,146). In this regard, the hypoxic storage may increase the generation of reactive oxygen species like  $H_2O_2$  because of the enhanced auto-oxidation of haemoglobin (with the formation of methaemoglobin) as well as the depletion of the cellular detoxifying mechanisms (147,148). Thus, it is likely that

ethanol may be oxidized to acetaldehyde either directly via  $H_2O_2$  and haeme (either as  $Fe^{3+}$  or  $Fe^{2+}$ ) as it occurs in the cytochrome (but with no further oxidation of acetaldehyde to acetic acid), or via a highly-reactive ferryl ( $Fe^{4+}$ ) haeme intermediate (94,96). The end of the catalysis would then depend on the consumption of the oxygen within the matrix rather than of haemoglobin. Theoretically, this should correspond to the loss of up to 0.2 g/L of ethanol according to the partial pressure of the oxygen in blood, a value that fit in the bias usually observed in long-term storage (96). It is clear then why freezing just slows down but does not stop the catalysis, venting the tube ignites new oxidative reactions *via* the residue haeme and the haemolysis before the storage prompts the oxidation *via* the premature depletion of the intracellular antioxidants.

#### Appendix C – the experimental design and the stability of blood alcohol concentration

Since the early studies on the preanalytics of BAC, the term "stability" has been used to indicate the change in ethanol concentration occurring between sample collection and analysis (51,58,60,61, 65,82,95,96,101,107,112,144). However, the exact attribution of a budget of bias to this term strictly depends on the experimental design by means of which it is measured the response of BAC under different conditions (139). In fact, the term "stability" should be used to indicate the bias arising from the spontaneous decay of ethanol within its matrix, and thus the amount loss just because of NEO (96). Accordingly, a "stabilized" specimen should be one where NEO was chemically inhibited (107).

By the point of view of the experimental design investigating the stability of BAC, the measured bias

is the omnibus response built up by factors like condition (i.e., temperature) and duration (i.e., time) of storage, and the interactions of all these terms (95). However, as NEO starts as soon as the whole blood is collected, the effect is easily confounded by the underlying sterility condition of the matrix, whose effect also interact with time and temperature. As a consequence, there arise different if not conflicting estimates of "stability" bias because of incomplete factorization of the design (i.e., incomplete attribution of the major variance components), particularly because of sterility issue (143,144,149-151). Otherwise, the "stability" bias arises as a trending (*i.e.*, systematically increasing) loss of ethanol, whilst other factors eventually acting as confounders (e.g., loss by leakage) show a rather random behaviour.